



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number <b>80-004945</b>	Instrument Location <b>Burleigh Mortar Detention</b>
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <b>Annual Inspection at Crime Laboratory</b>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P)
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **None Needed**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>5/15/23</b>
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date <b>22May2023</b>
--	------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer          Alcohol Analyzer  
North Dakota Model 8000          SN 80-004945  
Location = BMDC                  8164.14.00 09/16  
05/15/2023                          10:39

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:          Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?          No  
Disable On Memfull?      Yes  
# of Print Copies?        1  
Select Std (D/W/I)?      Dry  
Standard Value?          0.080  
Standard Lot #?          01923080A3  
Standard Cyl #?          55  
Standard Expiration?      02/05/2025  
Oper No?                  130315

Flow Cal. Date:            09/26/2012  
Slope                      622  
Intercept                 -485836

IR Calibration Date:      04/22/2016

	3um	9um
0th Coef(*100):	-23427	-26406
1st Coef(*100):	266989	144341
2nd Coef(*100):	2085	1078
H2O adj(mg/l*10k):	539	400

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JARED LEMIEUX

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = BMDC      8164.14.00 09/16  
05/15/2023      10:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:40
02 Std. Gas	0.082	10:40
03 Room Air	0.000	10:41
04 Std. Gas	0.082	10:41
05 Room Air	0.000	10:42
06 Std. Gas	0.082	10:42
07 Room Air	0.000	10:43

Lot No = 01923080A3  
Cyl No = 55  
Exp Date = 02/05/2025  
County = 08      Oper No. = 130315

  
Operator Signature  
JARED LEMIEUX

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = BMDC      8164.14.00 09/16  
05/15/2023      10:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:43
02 Std. Gas	RFI*	10:44
03 Room Air	RFI*	10:44

\*Invalid Test  
Inhibited - RFI

Lot No = 01923080A3  
Cyl No = 55  
Exp Date = 02/05/2025  
County = 08      Oper No. = 130315



---

Operator Signature  
JARED LEMIEUX

Remarks:

Form 106-I8000