




**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-005943</b>	Instrument Location <b>NCSO - NCHS hospital</b>
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <b>Annual Inspection at Crime Laboratory</b>	


Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	<b>4982</b>	Date <b>5/17/23</b>
--	-------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date <b>19May2023</b>
--	------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005943  
Location = NCSO                      8164.14.00 09/16  
05/17/2023                              10:54

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    34921080A1  
Standard Cyl #?                    26  
Standard Expiration?                02/05/2024  
Oper No?                            100754

Flow Cal. Date:                    06/20/2013  
Slope                                686  
Intercept                            -735246

IR Calibration Date:                08/26/2016  
                                        3um                                  9um

	3um	9um
0th Coef(*100):	-13589	-19547
1st Coef(*100):	284566	139614
2nd Coef(*100):	2717	1336
H2O adj(mg/l*10k):	381	404

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
DANIELLE BJORLIE

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005943  
Location = NCSO      8164.14.00 09/16  
05/17/2023      10:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:58
02 Std. Gas	0.078	10:59
03 Room Air	0.000	10:59
04 Std. Gas	0.079	11:00
05 Room Air	0.000	11:00
06 Std. Gas	0.079	11:01
07 Room Air	0.000	11:01

Lot No = 34921080A1  
Cyl No = 26  
Exp Date = 02/05/2024  
County = 32      Oper No. = 100754

  
\_\_\_\_\_  
Operator Signature  
DANIELLE BJORLIE

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005943  
Location = NCSO      8164.14.00 09/16  
05/17/2023      11:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	11:02
02 Room Air	0.000	11:02

\*Invalid Test  
Inhibited - RFI

Lot No = 34921080A1  
Cyl No = 26  
Exp Date = 02/05/2024  
County = 32      Oper No. = 100754

  
\_\_\_\_\_  
Operator Signature  
DANIELLE BJORLIE

Remarks:

Form 106-I8000