



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Cory Mortensen		
Location Cavalier		Intoxilyzer® Serial Number 80-006494
Gas Lot Number 34921080A1	Gas Cylinder Number 055	Gas Expiration Date 02/05/2024

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.079 AC
 B. 0.079 AC
 C. 0.079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature 	Date 5-18-2023
Reviewed By (Crime Laboratory Use Only) 	Date 19May2023

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = CAVA 8164.14.00 09/16
05/18/2023 16:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:31
02 Std. Gas	0.079	16:31
03 Room Air	0.000	16:32
04 Std. Gas	0.079	16:32
05 Room Air	0.000	16:33
06 Std. Gas	0.079	16:33
07 Room Air	0.000	16:34

Lot No = 34921080A1
Cyl No = 055
Exp Date = 02/05/2024
County = 34 Oper No. = 020832



Operator Signature
CORY MORTENSEN

Remarks: *ACA test*

Form 106-I8000