



**ETHANOL GAS STANDARD CYLINDER REPORT**  
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>David Stromberg</i>		
Location <i>Grand Forks County Corrections</i>	Intoxilyzer® Serial Number <i>80-003061</i>	
Gas Lot Number <i>01923080 AS</i>	Gas Cylinder Number <i>18</i>	Gas Expiration Date <i>07/05/2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.079* AC
  - B. *0.080* AC
  - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>5/18/2023</i>
Reviewed By (Crime Laboratory Use Only) <i>Aime Nahl</i>	Date 18May2023


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = GFSO      8164.14.00 09/16  
05/18/2023      10:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.079	10:58
03 Room Air	0.000	10:58
04 Std. Gas	0.080	10:59
05 Room Air	0.000	10:59
06 Std. Gas	0.080	10:59
07 Room Air	0.000	11:00

Lot No = 01923080A3  
Cyl No = 18  
Exp Date = 02/05/2025  
County = 18      Oper No. = 020836

  
Operator Signature  
DAVID STROMBERG

Remarks:

Form 106-I8000