



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-003061</b>	Instrument Location <b>Grand Forks County Corrections</b>
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <b>Annual Inspection at Crime Laboratory</b>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>5/18/2023</b>
-----------------------------------------------------------------------------------------------------------------	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <b>Aime Nahl</b>	Certified Date <b>18May2023</b>
-------------------------------------------	------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003061  
Location = GFSO                8164.14.00 09/16  
05/18/2023                      11:04

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            01923080A3  
Standard Cyl #?            18  
Standard Expiration?       02/05/2025  
Oper No?                    020836

Flow Cal. Date:            03/04/2011  
Slope                        700  
Intercept                   -623726

IR Calibration Date:       04/10/2015  
                                    3um                      9um

0th Coef(*100):	-25722	-17879
1st Coef(*100):	266163	136703
2nd Coef(*100):	2302	1151
H2O adj(mg/l*10k):	412	294

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
DAVID STROMBERG

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = GFSO      8164.14.00 09/16  
05/18/2023      11:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:06
02 Std. Gas	0.079	11:06
03 Room Air	0.000	11:07
04 Std. Gas	0.079	11:07
05 Room Air	0.000	11:07
06 Std. Gas	0.080	11:08
07 Room Air	0.000	11:08

Lot No = 01923080A3  
Cyl No = 18  
Exp Date = 02/05/2025  
County = 18      Oper No. = 020836

  
Operator Signature  
DAVID STROMBERG

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = GFSO      8164.14.00 09/16  
05/18/2023      11:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:15
02 Std. Gas	RFI*	11:16
03 Room Air	0.000	11:16

\*Invalid Test  
Inhibited - RFI

Lot No = 01923080A3  
Cyl No = 18  
Exp Date = 02/05/2025  
County = 18      Oper No. = 020836



---

Operator Signature  
DAVID STROMBERG

Remarks:

Form 106-I8000