



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4947 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or No (If Yes, Lot # 34921080A1 Cyl. # 40)
  - Keys  Yes or  No
  - Power Cord  Yes or  No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: DKPD
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring  Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or  MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 638 psi Regulator: 600 psi  
Display and Regulator ± 50 psi of each other  Yes or  No  
Gas tank tare necessary? Yes or  No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5318  
Lot #: 202201A  
Exp. Date: 11Jan24  
 Results ± 0.005 of known AC

|                                          |                                            |
|------------------------------------------|--------------------------------------------|
| Toxicology Unit - Breath Alcohol Section | Qualtrax ID: 11698 Revision: 1             |
| Intoxilyzer 8000 Annual Inspection       | Status: Published<br>Published: 03/24/2023 |
| UNCONTROLLED WHEN PRINTED                | Page 1 of 2                                |

2.  Configure simulator for the following test (Level 1,S).  
 Wet Calibration Check - High AC (Level 1,C)  
 Known Value  $\geq$  0.25 AC: 0.300 AC  
 Sim. Ser #: MP3067  
 Lot #: 202201F  
 Exp. Date: 18Jan24  
 Results  $\pm$  5% AC of known AC
3.  Configure dry gas standard for the remaining tests (Level 1,S).  
 Known Value : 0.080 AC  
 Gas Cylinder Lot #: 26021080A1  
 Cylinder #: 20  
 Exp. Date: 10/5/23
4.  Interferent Check (Level 1,B)  
 Known Value: 0.10 AC + 0.05% Acetone  
 Sim. Ser #: DR7345  
 Lot #: ICS 7  
 Exp. Date: N/A  
 Display reads "Interferent Detect"
5.  RFI Check (CMS Mode)  
 Display reads "RFI Detect"
6.  Dry Calibration Check (Level 1,C)  
 Test 1 0.080    Test 4 0.080    Test 7 0.080  
 Test 2 0.080    Test 5 0.079    Test 8 0.079  
 Test 3 0.080    Test 6 0.079    Test 9 0.080  
 Average 0.079  
 Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

---



---

Instrument is acceptable to be used in the field.  Yes or No

If No, state reason(s) why: \_\_\_\_\_

---

If Yes, change location code back to A.4.

[Signature]  
 Inspector Signature

01 June 2023  
 Date

[Signature]  
 Reviewer

01 June 2023  
 Date

|                                          |                                            |
|------------------------------------------|--------------------------------------------|
| Toxicology Unit - Breath Alcohol Section | Qualtrax ID: 11698 Revision: 1             |
| Intoxilyzer 8000 Annual Inspection       | Status: Published<br>Published: 03/24/2023 |
| UNCONTROLLED WHEN PRINTED                | Page 2 of 2                                |

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004947  
Location = TOXL                    8164.14.00 09/16  
06/01/2023                            08:05

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34921080A1  
Standard Cyl #?            040  
Standard Expiration?       02/05/2024  
Oper No?                    133237

Flow Cal. Date:            04/07/2020  
Slope                        618  
Intercept                   -565524

IR Calibration Date:        04/07/2020  
                                  3um                            9um

-----  
0th Coef(\*100):            -22371                        -26151  
1st Coef(\*100):            260822                        133081  
2nd Coef(\*100):            1936                            1144  
H2O adj(mg/l\*10k):        495                            535

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA NAREHOOD

Remarks:    *Print Test*








Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004947  
Location = TOXL                                      8164.14.00 09/16  
06/01/2023                                                                                      08:32

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Room Air      | 0.000 | 08:33 |
| 02 *Subject Test | INT*  | 08:33 |
| 03 Room Air      | 0.000 | 08:34 |

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                                      Weight = NA  
Test = OTH                                      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 26021080A1  
Cyl No = 20  
Expiration Date = 10/05/2023  
County = 08                                      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004947  
Location = TOXL                                      8164.14.00 09/16  
06/01/2023                                                                                      08:34

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Diagnostic    | OK    | 08:35 |
| 02 Room Air      | 0.000 | 08:36 |
| 03 *Subject Test | RFI*  | 08:36 |
| 04 Room Air      | 0.000 | 08:37 |

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                                      Weight = NA  
Test = OTH                                      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 26021080A1  
Cyl No = 20  
Expiration Date = 10/05/2023  
County = 08                                      Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: *RFI Check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.14.00 09/16  
06/01/2023      08:37

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 08:37 |
| 02 Std. Gas | 0.080 | 08:38 |
| 03 Room Air | 0.000 | 08:38 |
| 04 Std. Gas | 0.080 | 08:39 |
| 05 Room Air | 0.000 | 08:39 |
| 06 Std. Gas | 0.080 | 08:39 |
| 07 Room Air | 0.000 | 08:40 |

Lot No = 26021080A1  
Cyl No = 20  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dny cal check #1-3

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.14.00 09/16  
06/01/2023      08:40

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 08:41 |
| 02 Std. Gas | 0.080 | 08:41 |
| 03 Room Air | 0.000 | 08:42 |
| 04 Std. Gas | 0.079 | 08:42 |
| 05 Room Air | 0.000 | 08:43 |
| 06 Std. Gas | 0.079 | 08:43 |
| 07 Room Air | 0.000 | 08:44 |

Lot No = 26021080A1  
Cyl No = 20  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #4-4

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.14.00 09/16  
06/01/2023      08:44

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 08:44 |
| 02 Std. Gas | 0.080 | 08:45 |
| 03 Room Air | 0.000 | 08:45 |
| 04 Std. Gas | 0.079 | 08:45 |
| 05 Room Air | 0.000 | 08:46 |
| 06 Std. Gas | 0.080 | 08:46 |
| 07 Room Air | 0.000 | 08:47 |

Lot No = 26021080A1  
Cyl No = 20  
Exp Date = 10/05/2023  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks:

Dry cal check #7-9

Form 106-I8000