


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = OKPD 8164.16.00 09/18
09/25/2023 10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:19
02 Std. Gas	0.080	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.079	10:20
05 Room Air	0.000	10:20
06 Std. Gas	0.079	10:21
07 Room Air	0.000	10:21

Lot No = 14323080A1
Cyl No = 63
Exp Date = 06/05/2025
County = 11 Oper No. = 130690



Operator Signature
MATTHEW O'BRIEN

Remarks:

Form 106-I8000