



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>130690 (Matt O'Brien)</i>		
Location <i>Oakes PD</i>	Intoxilyzer® Serial Number <i>80-007095</i>	
Gas Lot Number <i>14323080A1</i>	Gas Cylinder Number <i>63</i>	Gas Expiration Date <i>6/05/2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.079* AC
 - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Matt O'Brien</i>	Date <i>9/15/23</i>
Reviewed By (Crime Laboratory Use Only) <i>Aime Wahl</i>	Date 25Sep2023


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = OKPD 8164.16.00 09/18
09/25/2023 10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:19
02 Std. Gas	0.080	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.079	10:20
05 Room Air	0.000	10:20
06 Std. Gas	0.079	10:21
07 Room Air	0.000	10:21

Lot No = 14323080A1
Cyl No = 63
Exp Date = 06/05/2025
County = 11 Oper No. = 130690



Operator Signature
MATTHEW O'BRIEN

Remarks:

Form 106-I8000