



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) ESTES, CHRIS		
Location DICKEY COUNTY COURTHOUSE		Intoxilyzer® Serial Number 80-005953
Gas Lot Number 14323080A1	Gas Cylinder Number 18	Gas Expiration Date 06/05/2025

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S)
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0 081).  
 A. 0.82 AC  
 B. 0.082AC  
 C. 0.082AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:  
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).  
 B. ACA Test Record.

*ALW  
10/15/23*

Chemical Test Operator Signature 	Date 10/5/2023
Reviewed By (Crime Laboratory Use Only) 	Date 10/5/2023

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58507

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005953  
Location = DCSO                    8164.14.00 09/16  
10/05/2023                            09:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:13
02 Std. Gas	0.082	09:14
03 Room Air	0.000	09:14
04 Std. Gas	0.082	09:15
05 Room Air	0.000	09:15
06 Std. Gas	0.082	09:15
07 Room Air	0.000	09:16

Lot No = 14323080A1  
Cyl No = 18  
Exp Date = 06/05/2025  
County = 11                            Oper No. = 020504

  
\_\_\_\_\_  
Operator Signature  
CHRIS ESTES

Remarks:

Form 106-I8000