



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL



CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Michelle Romans</i>		
Location <i>Williston</i>	Intoxilyzer® Serial Number <i>80-004944</i>	
Gas Lot Number <i>14323080A4</i>	Gas Cylinder Number <i>43</i>	Gas Expiration Date <i>6-5-2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.*082* AC
 - B. 0.*081* AC
 - C. 0.*082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature 	Date <i>10/12</i>
Reviewed By (Crime Laboratory Use Only) 	Date <i>13Oct2023</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WILL 8164.14.00 09/16
10/12/2023 19:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:09
02 Std. Gas	0.082	19:10
03 Room Air	0.000	19:10
04 Std. Gas	0.081	19:11
05 Room Air	0.000	19:11
06 Std. Gas	0.082	19:12
07 Room Air	0.000	19:12

Lot No = 14323080A4
Cyl No = 43
Exp Date = 06/05/2025
County = 53 Oper No. = 130739


Operator Signature
MICHELLE ROMANS

Remarks:

Form 106-I8000