

## ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM SFN 59282 (08/2018)

Chemical Test Operator Name (Print)	507	
Location Minat Police Pa		Intoxilyzer® Serial Number 80 - 006503
Gas Lot Number 1432 30 80 A 4	Gas Cylinder Number	Gas Expiration Date 66-05-2025

Check When Done:

Scan/Enter Gas Cylinder Information (Level 1, Function S).

2. Perform an ACA Test (Level 1, Function C). Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. 0. 12 \_ AC B. 0.12 \_ AC C. 0.12 \_ AC

Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

5. Send the Following to the Crime Laboratory:

A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature	Date
Dans 1019-	12.04-23
Reviewed By (Crime Laboratory Use/Only)	Date
	04Dec2023

Intoxilyzer Test Record and Checklist NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer North Dakota Model 8000 Location = MIPD 12/04/2023 Alcohol Analyzer SN 80-006503 8164.14.00 09/16 08:27

## DRY CAL CHECK

Test		AC	Time
01 Room 02 Std. 03 Room 04 Std. 05 Room 06 Std. 07 Room	Gas Air Gas Air Gas	0.000 0.082 0.000 0.082 0.000 0.082 0.000	08:28 08:28 08:28 08:29 08:29 08:30

Lot No = 14323080A4

Cyl No = 48

Exp Date = 06/05/2025

County = 51

Oper No. = 130907

Operator Signature JACOB OLSON

Remarks:

Form 106-I8000

The undersigned, having custody of the original public record, certifies that the information heron is a true and correct copy of the original document maintained as a part of the activities of this department.

417/7293

for

Minot Police

Office/Agency