



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Jeffrey D. Fleck</i>		
Location <i>LaMoure</i>		Intoxilyzer® Serial Number <i>80-006507</i>
Gas Lot Number <i>01923080A3</i>	Gas Cylinder Number <i>51</i>	Gas Expiration Date <i>02-05-2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- am* 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.082* AC
 - B. *0.082* AC
 - C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- am* 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Jeffrey D. Fleck</i>	Date <i>01/03/24</i>
Reviewed By (Crime Laboratory Use Only) <i>Christina</i>	Date <i>04 January 2024</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = LAMO 8164.14.00 09/16
01/03/2024 15:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:12
02 Std. Gas	0.082	15:13
03 Room Air	0.000	15:13
04 Std. Gas	0.082	15:13
05 Room Air	0.000	15:14
06 Std. Gas	0.082	15:14
07 Room Air	0.000	15:15

Lot No = 01923080A3
Cyl No = 51
Exp Date = 02/05/2025
County = 23 Oper No. = 010605


Operator Signature
JEFFREY FLECK

Remarks:

Form 106-I8000