## OFFICE OF ATTORNEY GENERAL CRIME LABORATORY DIVISION

Toxicology Section/Breath Program Chain of Custody Worksheet

Worksheet: BrW-012

## CHAIN OF CUSTODY BREATH ALCOHOL INSTRUMENT

Instrument: Maxilyzer 8000 Serial Number: 80-004948
Agency being assigned instrument: NDOAG-Cinnclab
Date IIJan2024
Name (print) Robert Walls  Signature Result Wall  Agency Walst lo. Strauge  Agency NOAG Come Lab.
Reason for transfer/remarks:  Instrument having temp. regulator issues. Swapped wt 80-004198
Or Shipped Via:
USPS UPS FedEx Other
Date Sending Instrument:// Air Bill No