



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Earl, William</i>		
Location <i>DEVL</i>	Intoxilyzer® Serial Number <i>80-06506</i>	
Gas Lot Number <i>14323080A4</i>	Gas Cylinder Number <i>004</i>	Gas Expiration Date <i>6/9/2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. *0.083* AC  
 B. *0.083* AC  
 C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i> <i>4534</i>	Date <i>1/25/2024</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>25Jan2024</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006506  
Location = DEVL      8164.14.00 09/16  
01/21/2024      04:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	04:15
02 Std. Gas	0.082	04:15
03 Room Air	0.000	04:16
04 Std. Gas	0.083	04:16
05 Room Air	0.000	04:17
06 Std. Gas	0.082	04:17
07 Room Air	0.000	04:18

Lot No = 14323080A4  
Cyl No = 4  
Exp Date = 06/05/2025  
County = 36      Oper No. = 132008



Operator Signature  
WILLIAM EARL

Remarks:

Form 106-I8000