



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Allen, Kathryn</i>		
Location <i>Cass County Jail</i>	Intoxilyzer® Serial Number <i>80-005944</i>	
Gas Lot Number <i>01923080A3</i>	Gas Cylinder Number <i>17</i>	Gas Expiration Date <i>2/5/2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. 0. *083* AC  
 B. 0. ~~082~~ AC  
 C. 0. *083* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Kathryn Allen 3671</i>	Date <i>1/22/2024</i>
Reviewed By (Crime Laboratory Use Only) <i>Aime Nahl</i>	Date <i>23Jan2024</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005944  
Location = CASS      8164.14.00 09/16  
01/22/2024      15:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:34
02 Std. Gas	0.083	15:35
03 Room Air	0.000	15:35
04 Std. Gas	0.082	15:36
05 Room Air	0.000	15:36
06 Std. Gas	0.083	15:37
07 Room Air	0.000	15:37

Lot No = 01923080A3  
Cyl No = 17  
Exp Date = 02/05/2025  
County = 09      Oper No. = 090801

  
Operator Signature  
KATHRYN ALLEN

Remarks:

Form 106-I8000

All checks within Normal Range!