



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Hiedeman, Thomas (</i>		
Location <i>Richland Co Jail; Wahpeton</i>	Intoxilyzer® Serial Number <i>80-006508</i>	
Gas Lot Number <i>28423080A3</i>	Gas Cylinder Number <i>47</i>	Gas Expiration Date <i>11/05/2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.081* AC
 B. *0.082* AC
 C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Thomas Hiedeman</i>	Date <i>01/31/2024</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>05Feb2024</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006508
Location = WAHP 8164.14.00 09/16
01/31/2024 01:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:01
02 Std. Gas	0.081	01:01
03 Room Air	0.000	01:02
04 Std. Gas	0.082	01:02
05 Room Air	0.000	01:03
06 Std. Gas	0.081	01:03
07 Room Air	0.000	01:04

Lot No = 28423080A3
Cyl No = 47
Exp Date = 11/05/2025
County = 39 Oper No. = 133155



Operator Signature
THOMAS HIEDEMAN

Remarks: ACA Test

Form 106-I8000