



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Mark Pearson</i>		
Location <i>Bismarck Cooling</i>	Intoxilyzer® Serial Number <i>80-004935</i>	
Gas Lot Number <i>14323080A4</i>	Gas Cylinder Number <i>39</i>	Gas Expiration Date <i>6/05/2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.079* AC
 B. *0.079* AC
 C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>Mark Pearson</i>	Date <i>2/1/24</i>
Reviewed By (Crime Laboratory Use Only) <i>Aime Wahl</i>	Date 05Feb2024

Intoxilyzer Test Record and Checklist
 ND096 Crime Lab. Div., Bismarck, ND 58501
 CHI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN 80-004935
 Location = LINT 8164.14.00 09/16
 02/01/2024 16:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:30
02 Std. Gas	0.079	16:30
03 Room Air	0.000	16:31
04 Std. Gas	0.079	16:31
05 Room Air	0.000	16:32
06 Std. Gas	0.079	16:32
07 Room Air	0.000	16:33

Lot No = 14323080A4
 Cyl No = 39
 Exp Date = 06/05/2025
 County = 15 Oper. No. = 130117
 Operator Signature
 MARK PEARSON

Remarks:

Form 106-18000