



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) CHRISTOPHER LEE ALMLIE		
Location Burke County Sheriff's Shop	Intoxilyzer® Serial Number 80-006673	
Gas Lot Number 14323080A4	Gas Cylinder Number 41	Gas Expiration Date 6-5-2025

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. **0.081** AC
 - B. **0.081** AC
 - C. **0.081** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature 	Date 2/5/2024
Reviewed By (Crime Laboratory Use Only) 	Date 06Feb2024

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
02/05/2024 17:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:36
02 Std. Gas	0.081	17:37
03 Room Air	0.000	17:37
04 Std. Gas	0.081	17:37
05 Room Air	0.000	17:38
06 Std. Gas	0.081	17:38
07 Room Air	0.000	17:39

Lot No = 14323080A4
Cyl No = 41
Exp Date = 06/05/2025
County = 07 Oper No. = 132664



Operator Signature
CHRISTOPHER ALMLIE

Remarks: *good*

Form 106-I8000