



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Kaitlyn Randall</i>		
Location <i>UNDP</i>	Intoxilyzer® Serial Number <i>80-005955</i>	
Gas Lot Number <i>28423080A.3</i>	Gas Cylinder Number <i>44</i>	Gas Expiration Date <i>11-05-2025</i>

Check When Done:

- KR-TLM* 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- KR-TLM* 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.081* AC
 - B. *0.080* AC
 - C. *0.080* AC
- KR* 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- KR* 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- KR* 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Kaitlyn Randall</i>	Date <i>2/5/2024</i>
Reviewed By (Crime Laboratory Use Only) <i>Ame Nahl</i>	Date <i>06Feb2024</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = UNDP 8164.14.00 09/16
02/05/2024 15:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:47
02 Std. Gas	0.081	15:47
03 Room Air	0.000	15:48
04 Std. Gas	0.080	15:48
05 Room Air	0.000	15:48
06 Std. Gas	0.080	15:49
07 Room Air	0.000	15:49

Lot No = 28423080A3
Cyl No = 44
Exp Date = 11/05/2025
County = 18 Oper No. = 132529

 #265
Operator Signature
KAITLYN RANDALL

Remarks:

Form 106-I8000