



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Rogers, Joseph L.		
Location Killdeer PD	Intoxilyzer® Serial Number 80-004938	
Gas Lot Number 14323080A4	Gas Cylinder Number 42	Gas Expiration Date 06/05/2025

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. 0.061 AC
  - B. 0.080 AC
  - C. 0.079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature 	Date 02/07/2024
Reviewed By (Crime Laboratory Use Only) 	Date 07Feb2024


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = KDPD      8164.14.00 09/16  
02/07/2024      07:54


DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:55
02 Std. Gas	0.081	07:55
03 Room Air	0.000	07:56
04 Std. Gas	0.080	07:56
05 Room Air	0.000	07:57
06 Std. Gas	0.079	07:57
07 Room Air	0.000	07:58

Lot No = 14323080A4  
Cyl No = 42  
Exp Date = 06/05/2025  
County = 13      Oper No. = 132016

  
Operator Signature  
JOSEPH ROGERS

Remarks:

True ar  
Date: 2/7/25  
Sign: 

Form 106-I8000