



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6684 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 14323080A1 Cyl. # 25)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: WFPD
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 488 psi Regulator: 500 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: NP3002
Lot #: 202301E
Exp. Date: 17Jan25
 Results ± 0.005 of known AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = TOXL 8164.14.00 09/16
04/02/2024 13:26

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.020
Standard Solution #? 202301E
Sim Serial #? MP3002
Oper No? 133237

Flow Cal. Date: 09/06/2016
Slope 709
Intercept -725441

IR Calibration Date: 09/06/2016
 3um 9um

0th Coef(*100): -16013 -23467
1st Coef(*100): 256431 135731
2nd Coef(*100): 2904 1268
H2O adj(mg/l*10k): 553 514

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = TOXL 8164.14.00 09/16
04/02/2024 13:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Sol.	0.020	13:30
03 Room Air	0.000	13:31
04 Std. Sol.	0.020	13:31
05 Room Air	0.000	13:32
06 Std. Sol.	0.020	13:33
07 Room Air	0.000	13:33

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC check - 0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = TOXL 8164.14.00 09/16
04/02/2024 13:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:37
02 Std. Sol.	0.390	13:37
03 Room Air	0.000	13:38
04 Std. Sol.	0.400	13:39
05 Room Air	0.000	13:39
06 Std. Sol.	0.402	13:40
07 Room Air	0.000	13:40

08 Sim Temp = 34.0°C

Simul Ser No = MP6036
Std Sol No = 202212C
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac Check - 0.400Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = TOXL 8164.14.00 09/16
04/02/2024 13:43

Test	AC	Time
01 Room Air	0.000	13:44
02 *Subject Test	INT*	13:44
03 Room Air	0.000	13:45

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 39
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = TOXL 8164.14.00 09/16
04/02/2024 13:46

Test	AC	Time
01 Diagnostic	OK	13:47
02 Room Air	0.000	13:47
03 *Subject Test	RFI*	13:48
04 Room Air	0.000	13:48

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 39
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = TOXL 8164.14.00 09/16
04/02/2024 13:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:49
02 Std. Gas	0.082	13:49
03 Room Air	0.000	13:50
04 Std. Gas	0.081	13:50
05 Room Air	0.000	13:50
06 Std. Gas	0.081	13:51
07 Room Air	0.000	13:51

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry Cal check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = TOXL 8164.14.00 09/16
04/02/2024 13:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:52
02 Std. Gas	0.081	13:53
03 Room Air	0.000	13:53
04 Std. Gas	0.080	13:53
05 Room Air	0.000	13:54
06 Std. Gas	0.081	13:54
07 Room Air	0.000	13:55

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = TOXL 8164.14.00 09/16
04/02/2024 13:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Gas	0.080	13:56
03 Room Air	0.000	13:57
04 Std. Gas	0.081	13:57
05 Room Air	0.000	13:58
06 Std. Gas	0.080	13:58
07 Room Air	0.000	13:59

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000