



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number <u>00-004940</u>	Instrument Location <u>Watford city AEN</u>
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Annual Inspection at Crime Laboratory</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN 50496 Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <u>[Signature]</u>	Date <u>04/02/24</u>
--	----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <u>[Signature]</u>	Certified Date <u>29 April 2024</u>
--	-------------------------------------

- Did not have correct location code.
 Reached out to have location code updated and new print test was sent. - AEN

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/02/2024 12:39

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 18
Standard Expiration? 06/05/2025
Oper No? 132007

Flow Cal. Date: 08/28/2018
Slope 642
Intercept -606282

IR Calibration Date: 04/01/2024
3um 9um

0th Coef(*100): -23587 -22037
1st Coef(*100): 256489 131821
2nd Coef(*100): 3363 1546
H2O adj(mg/l*10k): 339 298

***** Printer Test End *****

Kari Stuart #310
Operator Signature
KARI STUART

Remarks:

STATE OF NORTH DAKOTA } ss
COUNTY OF MCKENZIE }
I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Bismarck City, North Dakota.
on this and day of April 2024
Kari Stuart #310
Bismarck City Police Department

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/03/2024 06:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	06:05
02 Std. Gas	0.078	06:06
03 Room Air	0.000	06:06
04 Std. Gas	0.078	06:07
05 Room Air	0.000	06:07
06 Std. Gas	0.078	06:07
07 Room Air	0.000	06:08

Lot No = 14323080A4
Cyl No = 18
Exp Date = 06/05/2025
County = 27 Oper No. = 132007

Kari Stuart 3410

Operator Signature
KARI STUART

Remarks:

ACA

Form 106-I8000

STATE OF NORTH DAKOTA } SS
COUNTY OF MCKENZIE
I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.
on this 3rd day of April 2024
Kari Stuart
Watford City Police Department

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.14.00 09/16
04/03/2024 06:09

Test	AC	Time
01 Room Air	RFI*	06:10
02 Room Air	RFI*	06:10

*Invalid Test
Inhibited - RFI

Sub Name = STUART, KARI A
Sub DOB = 09/09/1990
Sub Sex = Female Weight = 157
Test = OTH Cit = 00000000
Dr. Lic. = ND/STU900000
Lot No = 14323080A4
Cyl No = 18
Expiration Date = 06/05/2025
County = 27 Oper No. = 132007



Operator Signature
KARI STUART

Remarks:

RFI

Form 106-I8000

STATE OF NORTH DAKOTA } ss
COUNTY OF MCKENZIE

I certify that the foregoing is a true and correct
copy of the original document on file in this office.
Dated at Watford City, North Dakota.

on this 3rd day of April 2024


Watford City Police Department

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = WCPD 8164.14.00 09/16
04/09/2024 18:26

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvWxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvWxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 18
Standard Expiration? 06/05/2025
Oper No? 133050

Flow Cal. Date: 08/28/2018
Slope 642
Intercept -606282

IR Calibration Date: 04/01/2024
 3um 9um

	3um	9um
0th Coef(*100):	-23587	-22037
1st Coef(*100):	256489	131821
2nd Coef(*100):	3363	1546
H2O adj(mg/l*10k):	339	298

***** Printer Test End *****



Operator Signature
MATTHEW EARL

Remarks: *print test*

Form 106-I8000