



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6688 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 28423080A3 Cyl. # 44)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: UNDP
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 971 psi Regulator: 975 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002
Lot #: 202301E
Exp. Date: 17Jan25
 Results ± 0.005 of known AC

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.400 AC
 Sim. Ser #: MP6036
 Lot #: 202212C
 Exp. Date: 08Dec24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 14323080A4
 Cylinder #: 13
 Exp. Date: 6/5/25
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR7344
 Lot #: ICS 8
 Exp. Date: NIA
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.082</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: ~~HHA~~ Initial dry calibration checks were high. Calibration adjustment performed and the low AC check, high AC check and all the dry calibration checks were reran.

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4.

An Mel
 Inspector Signature

04Apr2024
 Date

Janelle Putschner
 Reviewer

05 Apr 2024
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 08:33

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 28423080A3
Standard Cyl #? 44
Standard Expiration? 11/05/2025
Oper No? 133237

Flow Cal. Date: 09/07/2016
Slope 698
Intercept -642832

IR Calibration Date: 09/07/2016
 3um 9um

0th Coef(*100): -10822 -23152
1st Coef(*100): 262468 134492
2nd Coef(*100): 3796 1436
H2O adj(mg/l*10k): 665 568

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 09:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:11
02 Std. Sol.	0.022	09:12
03 Room Air	0.000	09:12
04 Std. Sol.	0.022	09:13
05 Room Air	0.000	09:13
06 Std. Sol.	0.022	09:14
07 Room Air	0.000	09:15

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC check - 0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 09:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:18
02 Std. Sol.	0.392	09:20
03 Room Air	0.000	09:20
04 Std. Sol.	0.397	09:21
05 Room Air	0.000	09:22
06 Std. Sol.	0.399	09:22
07 Room Air	0.000	09:23

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac Check - 0.400Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 09:29

Test	AC	Time
01 Room Air	0.000	09:30
02 *Subject Test	INT*	09:31
03 Room Air	0.000	09:32

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 13
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 09:32

Test	AC	Time
01 Diagnostic	OK	09:33
02 Room Air	0.000	09:33
03 *Subject Test	RFI*	09:34
04 Room Air	0.000	09:34

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 13
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 09:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:35
02 Std. Gas	0.082	09:35
03 Room Air	0.000	09:36
04 Std. Gas	0.083	09:36
05 Room Air	0.000	09:37
06 Std. Gas	0.082	09:37
07 Room Air	0.000	09:38

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 09:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:39
02 Std. Gas	0.083	09:39
03 Room Air	0.000	09:39
04 Std. Gas	0.083	09:40
05 Room Air	0.000	09:40
06 Std. Gas	0.083	09:41
07 Room Air	0.000	09:41

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check
needs cal adj.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 10:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:59
02 Std. Sol.	0.020	11:00
03 Room Air	0.000	11:01
04 Std. Sol.	0.020	11:01
05 Room Air	0.000	11:02
06 Std. Sol.	0.021	11:02
07 Room Air	0.000	11:03

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks:

~~Dry cal che~~
Low Ac Check-0.020 AC
post cal. adj.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 11:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:05
02 Std. Sol.	0.394	11:05
03 Room Air	0.000	11:06
04 Std. Sol.	0.397	11:07
05 Room Air	0.000	11:07
06 Std. Sol.	0.398	11:08
07 Room Air	0.000	11:08

08 Sim Temp = 34.0°C

Simul Ser No = MP6036
Std Sol No = 202212C
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check- 0.400AC
post cal adj.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 11:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:12
02 Std. Gas	0.081	11:12
03 Room Air	0.000	11:13
04 Std. Gas	0.081	11:13
05 Room Air	0.000	11:14
06 Std. Gas	0.081	11:14
07 Room Air	0.000	11:15

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3
post cal adj.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 11:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:16
02 Std. Gas	0.081	11:16
03 Room Air	0.000	11:16
04 Std. Gas	0.081	11:17
05 Room Air	0.000	11:17
06 Std. Gas	0.081	11:18
07 Room Air	0.000	11:18

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6
post cal adj.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = TOXL 8164.14.00 09/16
04/04/2024 11:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:19
02 Std. Gas	0.082	11:20
03 Room Air	0.000	11:20
04 Std. Gas	0.081	11:20
05 Room Air	0.000	11:21
06 Std. Gas	0.082	11:21
07 Room Air	0.000	11:22

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9
post cal adj.

Form 106-I8000