



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 7097 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 14323080 A4 Cyl. # 10)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: ROLL
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 928 psi Regulator: 950 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002
Lot #: 202301E
Exp. Date: 17Jan25
 Results ± 0.005 of known AC

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.400 AC
 Sim. Ser #: MP6036
 Lot #: 202212C
 Exp. Date: 08Dec24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 28423080A3
 Cylinder #: 39
 Exp. Date: 11/5/25
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR7344
 Lot #: ICS 8
 Exp. Date: NIA
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)

Test 1	<u>0.079</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
		Average		<u>0.079</u>	

 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: Initial dry calibration check was (0.083, 0.083, 0.083). Calibration adjustment performed. Low AC, high AC and dry calibration checks were ran post-cal adjustment.

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4. NIA Instrument swapped.

[Signature]
 Inspector Signature

04Apr2024
 Date

[Signature]
 Reviewer

05Apr2024
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/03/2024 12:19

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 10
Standard Expiration? 06/05/2025
Oper No? 133237

Flow Cal. Date: 09/03/2020
Slope 660
Intercept -636264

IR Calibration Date: 09/03/2020
 3um 9um

0th Coef(*100): -36810 -25919
1st Coef(*100): 254044 139068
2nd Coef(*100): 1858 1258
H2O adj(mg/l*10k): 903 584

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/03/2024 12:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:40
02 Std. Sol.	0.020	12:41
03 Room Air	0.000	12:42
04 Std. Sol.	0.020	12:42
05 Room Air	0.000	12:43
06 Std. Sol.	0.020	12:44
07 Room Air	0.000	12:44

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = 202301E
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC Check -0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/03/2024 12:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:52
02 Std. Sol.	0.397	12:52
03 Room Air	0.000	12:53
04 Std. Sol.	0.397	12:54
05 Room Air	0.000	12:54
06 Std. Sol.	0.397	12:55
07 Room Air	0.000	12:56


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

High AC Check - 0.400AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/03/2024 12:57

Test	AC	Time
01 Room Air	0.000	12:58
02 *Subject Test	INT*	12:58
03 Room Air	0.000	12:59

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 39
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent Check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/03/2024 12:59

Test	AC	Time
01 Diagnostic	OK	13:00
02 Room Air	0.000	13:01
03 *Subject Test	RFI*	13:01
04 Room Air	0.000	13:01

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 39
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: *RFI check*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/03/2024 14:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:33
02 Std. Gas	0.083	14:33
03 Room Air	0.000	14:34
04 Std. Gas	0.083	14:34
05 Room Air	0.000	14:34
06 Std. Gas	0.083	14:35
07 Room Air	0.000	14:35

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check
needs cal adj.*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/04/2024 10:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Sol.	0.020	10:08
03 Room Air	0.000	10:08
04 Std. Sol.	0.020	10:09
05 Room Air	0.000	10:09
06 Std. Sol.	0.020	10:10
07 Room Air	0.000	10:11

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC CHECK - 0.020AC
post cal adj*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/04/2024 10:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:12
02 Std. Sol.	0.395	10:13
03 Room Air	0.000	10:13
04 Std. Sol.	0.397	10:14
05 Room Air	0.000	10:15
06 Std. Sol.	0.397	10:15
07 Room Air	0.000	10:16

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.400 AC
post cal. adj.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/04/2024 10:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:28
02 Std. Gas	0.079	10:29
03 Room Air	0.000	10:29
04 Std. Gas	0.079	10:29
05 Room Air	0.000	10:30
06 Std. Gas	0.080	10:30
07 Room Air	0.000	10:31

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3
post cal. adj.

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/04/2024 10:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:33
02 Std. Gas	0.079	10:33
03 Room Air	0.000	10:34
04 Std. Gas	0.079	10:34
05 Room Air	0.000	10:35
06 Std. Gas	0.079	10:35
07 Room Air	0.000	10:35

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6
post cal adj.

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
04/04/2024 10:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.079	10:37
03 Room Air	0.000	10:37
04 Std. Gas	0.080	10:38
05 Room Air	0.000	10:38
06 Std. Gas	0.079	10:39
07 Room Air	0.000	10:39

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check #7-9
post cal adj.*

Form 106-I8000