



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Zachary T Smith</i>		
Location <i>Rolla</i>	Intoxilyzer® Serial Number <i>80-004954</i>	
Gas Lot Number <i>14323080A4</i>	Gas Cylinder Number <i>10</i>	Gas Expiration Date <i>6/5/2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. 0. 081 AC
B. 0. 081 AC
C. 0. 081 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Zachary T Smith</i>	Date <i>4/4/2024</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>17 Apr 2024</i>

Incorrect location code originally. Contacted Zachary Smith. Location code was changed and a new print test was sent to show correct location code. AEN 17Apr2024

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
04/04/2024 14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.081	14:38
03 Room Air	0.000	14:38
04 Std. Gas	0.081	14:38
05 Room Air	0.000	14:39
06 Std. Gas	0.081	14:39
07 Room Air	0.000	14:40

Lot No = 14323080A4
Cyl No = 10
Exp Date = 06/05/2025
County = 40 Oper No. = 131486

Zachary Smith
Operator Signature
ZACHARY SMITH

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.14.00 09/16
04/09/2024 15:23

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 10
Standard Expiration? 06/05/2025
Oper No? 131486

Flow Cal. Date: 08/27/2012
Slope 670
Intercept -570610

IR Calibration Date: 04/02/2024
 3um 9um

	3um	9um
0th Coef(*100):	-20499	-21878
1st Coef(*100):	255404	129613
2nd Coef(*100):	1977	1285
H2O adj(mg/l*10k):	414	417

***** Printer Test End *****



Operator Signature
ZACHARY SMITH

Remarks: