



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-004954	Instrument Location Bolla
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Annual Inspection at Crime Laboratory</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature Zachary Smith	Date 4/4/24
---	-----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By [Signature]	Certified Date 17 Apr 2024
---	--------------------------------------

Incorrect location code originally. Contacted Zachary Smith. Location code was changed and new print test was sent to show location code correction. AEN 17 Apr 2024

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
04/04/2024 14:32

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 10
Standard Expiration? 06/05/2025
Oper No? 131486

Flow Cal. Date: 08/27/2012
Slope 670
Intercept -570610

IR Calibration Date: 04/02/2024

	3um	9um
--	-----	-----

0th Coef(*100):	-20499	-21878
1st Coef(*100):	255404	129613
2nd Coef(*100):	1977	1285
H2O adj(mg/l*10k):	414	417

***** Printer Test End *****

Zachary Smith

Operator Signature
ZACHARY SMITH

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
04/04/2024 14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.081	14:38
03 Room Air	0.000	14:38
04 Std. Gas	0.081	14:38
05 Room Air	0.000	14:39
06 Std. Gas	0.081	14:39
07 Room Air	0.000	14:40

Lot No = 14323080A4
Cyl No = 10
Exp Date = 06/05/2025
County = 40 Oper No. = 131486

Zachary Smith
Operator Signature
ZACHARY SMITH

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.14.00 09/16
04/04/2024 14:42

Test	AC	Time
01 Room Air	RFI*	14:45
02 Room Air	RFI*	14:45

*Invalid Test
 Inhibited - RFI

Sub Name = RFI TEST, RFI TEST RFI TEST
Sub DOB = 09/02/1995
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 14323080A4
Cyl No = 10
Expiration Date = 06/05/2025
County = 40 Oper No. = 131486

Zachary Smith

Operator Signature
ZACHARY SMITH

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.14.00 09/16
04/09/2024 15:23

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 10
Standard Expiration? 06/05/2025
Oper No? 131486

Flow Cal. Date: 08/27/2012
Slope 670
Intercept -570610

IR Calibration Date: 04/02/2024
 3um 9um

0th Coef(*100): -20499 -21878
1st Coef(*100): 255404 129613
2nd Coef(*100): 1977 1285
H2O adj(mg/l*10k): 414 417

***** Printer Test End *****



Operator Signature
ZACHARY SMITH

Remarks: