



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4189 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:  
Gas Cylinder Yes or No (If Yes, Lot # \_\_\_\_\_ Cyl. # \_\_\_\_\_)  
Keys Yes or No  
Power Cord Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: TOXL
5.  Battery Check  
Was the external battery pack replaced? Yes or No
6.  O-Rings  
Replaced Simulator O-Ring Yes or No  
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 738 psi Regulator: 750 psi  
Display and Regulator  $\pm$  50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value  $\leq$  0.03 AC: 0.020 AC  
Sim. Ser #: MP3002  
Lot #: 202301E  
Exp. Date: 17Jan25  
 Results  $\pm$  0.005 of known AC

2.  Configure simulator for the following test (Level 1,S).  
 Wet Calibration Check - High AC (Level 1,C)  
 Known Value  $\geq$  0.25 AC: 0.400 AC  
 Sim. Ser #: MP6036  
 Lot #: 202212C  
 Exp. Date: 08 Dec 24  
 Results  $\pm$  5% AC of known AC
3.  Configure dry gas standard for the remaining tests (Level 1,S).  
 Known Value : 0.080 AC  
 Gas Cylinder Lot #: 28423080A3  
 Cylinder #: 37  
 Exp. Date: 11/05/25
4.  Interferent Check (Level 1,B)  
 Known Value: 0.10 AC + 0.05% Acetone  
 Sim. Ser #: DR7344  
 Lot #: ICS 8  
 Exp. Date: NIA  
 Display reads "Interferent Detect"
5.  RFI Check (CMS Mode)  
 Display reads "RFI Detect"
6.  Dry Calibration Check (Level 1,C)  

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.082</u>	Test 6	<u>0.082</u>	Test 9	<u>0.081</u>
		Average	<u>0.081</u>		

 Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: NIA

---



---

Instrument is acceptable to be used in the field.  Yes or No  
 If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4.  NIA

[Signature]  
 Inspector Signature

05 Apr 2024  
 Date

[Signature]  
 Reviewer

08 Apr 2024  
 Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = TOXL      8164.14.00 09/16  
04/05/2024      09:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:32
02 Std. Sol.	0.023	09:33
03 Room Air	0.000	09:33
04 Std. Sol.	0.022	09:34
05 Room Air	0.000	09:35
06 Std. Sol.	0.022	09:35
07 Room Air	0.000	09:36

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: ~~High AC Check - 0.400AC~~  
LOW      0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = TOXL      8164.14.00 09/16  
04/05/2024      09:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:38
02 Std. Sol.	0.396	09:38
03 Room Air	0.000	09:39
04 Std. Sol.	0.399	09:40
05 Room Air	0.000	09:40
06 Std. Sol.	0.400	09:41
07 Room Air	0.000	09:42

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.400 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = TOXL      8164.14.00 09/16  
04/05/2024      09:44

Test	AC	Time
01 Room Air	0.000	09:45
02 *Subject Test	INT*	09:46
03 Room Air	0.000	09:47

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 28423080A3  
Cyl No = 37  
Expiration Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = TOXL      8164.14.00 09/16  
04/05/2024      09:47

Test	AC	Time
01 Diagnostic	OK	09:48
02 Room Air	0.000	09:49
03 *Subject Test	RFI*	09:49
04 Room Air	0.000	09:50

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 28423080A3  
Cyl No = 37  
Expiration Date = 11/05/2025  
County = 08      Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: *RFI check*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = TOXL      8164.14.00 09/16  
04/05/2024      09:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:50
02 Std. Gas	0.081	09:51
03 Room Air	0.000	09:51
04 Std. Gas	0.081	09:52
05 Room Air	0.000	09:52
06 Std. Gas	0.082	09:52
07 Room Air	0.000	09:53

Lot No = 28423080A3  
Cyl No = 37  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = TOXL      8164.14.00 09/16  
04/05/2024      09:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:55
02 Std. Gas	0.081	09:55
03 Room Air	0.000	09:56
04 Std. Gas	0.081	09:56
05 Room Air	0.000	09:56
06 Std. Gas	0.082	09:57
07 Room Air	0.000	09:57

Lot No = 28423080A3  
Cyl No = 37  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #4-4

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = TOXL      8164.14.00 09/16  
04/05/2024      09:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:59
02 Std. Gas	0.081	09:59
03 Room Air	0.000	10:00
04 Std. Gas	0.081	10:00
05 Room Air	0.000	10:01
06 Std. Gas	0.081	10:01
07 Room Air	0.000	10:02

Lot No = 28423080A3  
Cyl No = 37  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000