

INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM SFN 59281 (06/2018)

Serial Number 80 - 006 490	Instrument Location	
Reason for Install/Repair	17,577	
☐ Install After Receiving From Crime Laboratory ☐ Install A	fter Location Change	
M Other (Specify) Onnual Inspection at	- Crime Laboratory	
	0	
Check When Done:		
Surge Protector Installed/Property Grounded.		
2. Telephone Line Connected to Intoxilyzer® 8000.		
3. Breath Tube Heated.		
4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Loc	ation; Level 2, Function E).	
5. Scan/Enter Gas Cylinder Information (Level 1, Function S).	
6. Run Tests:		
A. Print Test (Level1, Function P).		
B. ACA Test (Level 1, Function C).		
C. Radio Frequency Interference (RFI) Test (CMS Mode	or Level 1, Function B or C; Key Radio Du	uring Test).
7. Repair and/or Maintenance Performed (if any):		
8. Complete the Top Portion of the Intoxilyzer® Record (SF	N50496, Form 120-G) and Place it by the l	ntoxilyzer® for Use.
9. File Previous Intoxilyzer® Record (SFN504096, Form 120		
10. Send the Following to the Crime Laboratory: Completed 104-G), Print Test, ACA Test, and RFI Test.	Intoxilyzer® 8000 Installation and Repair (Checkout (SFN59281, Form
Field Inspector Signature		Date
		04/08/2024
Crime Laboratory Use Only		
This installation has been reviewed and the instrument is approved t	o be used for the analysis of breath to dete	ermine alcohol concentration from
the date the Field Inspector performed the installation. This record of County of Burleigh, North Dakota, is certified to be a true and correct	on file at the Office of Attorney General, Cr t copy of the documents received.	ime Laboratory Division, in the
AA	1-A-	Certified Date
Reviewed/Certified By		OOA pril 2024

Intoxilyzer Test Record and Checklist NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = ASHL 8164.14.00 09/16
04/08/2024 11:21

********** Printer Test *********

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=| ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

Current Instrument Setup Data Entry Mode: Enabled Start Test Sequence: DABACABA Display Prelim Rslt? Yes Display Third Digit? Yes Inhib Printer(Y/N)? No Display Volume? No Disable On Memfull? Yes # of Print Copies? 1 Select Std (D/W/I)? Dry Standard Value? 0.080 Standard Lot #? Standard Cyl #? 01923080A3 62 Standard Expiration? 02/05/2025 Oper No? 132466

Flow Cal. Date: 09/04/2020 Slope 649 Intercept -598817

IR Calibration Date: 09/04/2020

3um 9um

Oth Coef(*100): -25190 -16392
1st Coef(*100): 271645 136906
2nd Coef(*100): 2533 1222

****** Printer Test End *******

309

Operator Signature ALEXANDER WEHRUNG

Remarks: Print test

H2O adj(mg/l*10k): 460

Form 106-I8000

Intoxilyzer Test Record and Checklist NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer
North Dakota Model 8000
Location = ASHL
04/08/2024

Alcohol Analyzer SN 80-006490 8164.14.00 09/16 12:48

DRY CAL CHECK

Т	est	AC	Time
01	Room Air	0.000	12:49
02	Std. Gas	0.079	12:49
03	Room Air	0.000	12:49
04	Std. Gas	0.079	12:49
05	Room Air	0.000	MANUAL 12.032-10-1
06	Std. Gas	0.079	12:50
	Room Air	0.000	12:51
		0.000	12:51

Lot No = 01923080A3

Cyl No = 62

Exp Date = 02/05/2025

County = 26

Oper No. = 132466

Operator Signature ALEXANDER WEHRUNG

Remarks: ACA Test

Form 106-I8000

Intoxilyzer Test Record and Checklist NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer North Dakota Model 8000 Location = ASHL 04/08/2024

Alcohol Analyzer SN 80-006490 8164.14.00 09/16 12:52

Test	AC	Time
01 Room Air	RFI*	12:53
02 Room Air	0.000	12:54

*Invalid Test Inhibited - RFI

Sub Name = TEST, TEST T
Sub DOB = 04/08/1998
Sub Sex = Male

Test = DUI

Dr. Lic. = ND/TES987756
Lot No = 01923080A3
Cyl No = 62

Weight = 300
Cit = 026

Operator Signature ALEXANDER WEHRUNG

Remarks: RFI Fest

Form 106-I8000