



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4942 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: TOXL
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
 - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
 - Display: 136 psi Regulator: 150 psi
 - Display and Regulator ± 50 psi of each other Yes or No
 - Gas tank tare necessary? Yes or No
 - If Yes, display readings after tare (Level 3,M,C,G):
 - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
 - Wet Calibration Check - Low AC (Level 1,C)
 - Known Value ≤ 0.03 AC: 0.020 AC
 - Sim. Ser #: MP3002
 - Lot #: 202301E
 - Exp. Date: 17Jan25
 - Results ± 0.005 of known AC

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.400 AC
 Sim. Ser #: MP6036
 Lot #: 202212C
 Exp. Date: 08 Dec 24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 28423080A3
 Cylinder #: 39
 Exp. Date: 11/5/25
4. Interferent Check (Level 1,B)
 Known Value: 0.16 AC + 0.05% Acetone
 Sim. Ser #: DR7344
 Lot #: ICS 8
 Exp. Date: N/A
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.079</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.079</u>
Test 3	<u>0.080</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
		Average	<u>0.080</u>		

 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: Initial dry cal. check was high (.083, .082, .083). Calibration adjustment was performed. Low AC, High AC and all dry calibration checks were ran post-cal. adj.

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4. N/A

[Signature]
 Inspector Signature

09 Apr 2024
 Date

[Signature]
 Reviewer

10 Apr 2024
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 09:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:43
02 Std. Sol.	0.020	09:44
03 Room Air	0.000	09:44
04 Std. Sol.	0.021	09:45
05 Room Air	0.000	09:45
06 Std. Sol.	0.021	09:46
07 Room Air	0.000	09:47

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC check - 0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 10:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:01
02 Std. Sol.	0.409	10:02
03 Room Air	0.000	10:02
04 Std. Sol.	0.409	10:03
05 Room Air	0.000	10:04
06 Std. Sol.	0.409	10:04
07 Room Air	0.000	10:05

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.400AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 10:07

Test	AC	Time
01 Room Air	0.000	10:07
02 *Subject Test	INT*	10:08
03 Room Air	0.000	10:09

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 39
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 10:09

Test	AC	Time
01 Diagnostic	OK	10:10
02 Room Air	0.000	10:11
03 *Subject Test	RFI*	10:12
04 Room Air	0.000	10:12

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 28423080A3
Cyl No = 39
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 10:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:13
02 Std. Gas	0.083	10:14
03 Room Air	0.000	10:14
04 Std. Gas	0.082	10:14
05 Room Air	0.000	10:15
06 Std. Gas	0.083	10:15
07 Room Air	0.000	10:16

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check
needs cal. adj.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 11:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:43
02 Std. Sol.	0.018	11:44
03 Room Air	0.000	11:45
04 Std. Sol.	0.019	11:46
05 Room Air	0.000	11:46
06 Std. Sol.	0.019	11:47
07 Room Air	0.000	11:47

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = 202301E
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC CHECK - 0.020AC
post cal adj.*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 11:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Sol.	0.403	11:56
03 Room Air	0.000	11:56
04 Std. Sol.	0.403	11:57
05 Room Air	0.000	11:57
06 Std. Sol.	0.401	11:58
07 Room Air	0.000	11:59

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.400 AC
post cal adj.

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 12:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:01
02 Std. Gas	0.080	12:01
03 Room Air	0.000	12:02
04 Std. Gas	0.080	12:02
05 Room Air	0.000	12:03
06 Std. Gas	0.080	12:03
07 Room Air	0.000	12:04

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3
post cal adj.

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 12:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:06
02 Std. Gas	0.080	12:07
03 Room Air	0.000	12:07
04 Std. Gas	0.080	12:08
05 Room Air	0.000	12:08
06 Std. Gas	0.081	12:08
07 Room Air	0.000	12:09

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6
post cal adj.
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = TOXL 8164.14.00 09/16
04/09/2024 12:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:10
02 Std. Gas	0.079	12:10
03 Room Air	0.000	12:11
04 Std. Gas	0.079	12:11
05 Room Air	0.000	12:11
06 Std. Gas	0.081	12:12
07 Room Air	0.000	12:12

Lot No = 28423080A3
Cyl No = 39
Exp Date = 11/05/2025
County = 08 Oper No. = 133237

Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9
post cal adj.

Form 106-I8000