



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 REPAIR AND MAINTENANCE

Intoxilyzer® 8000 Serial Number: 80-00 4194 Maintenance Location: ToxL

1. Reported Issue and Issue Found:

High calibration values reported.

2. Repair/Maintenance Performed:

Calibration adjusted

3. Does the instrument flow sensor need to be adjusted? Yes or No

- If Yes, complete Flow Sensor Adjustment and Calibration (Document ID: 11872), then go to step 4.
- If No, go to step 4.

4. Does the instrument calibration need to be adjusted? Yes or No

- If Yes, complete Intoxilyzer® 8000 Calibration Adjustment (Document ID: 11859), then go to step 5.
- If No, go to step 5.

5. Was the instrument cover removed for the repair? Yes or No

- If Yes, complete Intoxilyzer® 8000 Initial Inspection (Document ID: 11784), then go to step 6.
- If No, go to step 6.

6. Is an Annual Inspection due on the Intoxilyzer® 8000? Yes or No

- If Yes, complete Intoxilyzer® 8000 Annual Inspection (Document ID: 11698), step 7 is not needed, and sign below.
- If No, go to step 7.


7. Was an Intoxilyzer® 8000 Initial Inspection (Document ID: 11784) completed in step 5? Yes or No

- If Yes, no further action is needed and sign below.
- If No, Complete Intoxilyzer® 8000 Calibration (Document ID: 11871) and sign below.

Remarks/Notes: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why:


Breath Alcohol Analyst Signature

10 Apr 2024
Date


Reviewer Signature

11 Apr 2024
Date