



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 0500 Inspection Location: ToxL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # 01923080A3 Cyl. # 59)
Keys Yes or No
Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: HARV
5. Battery Check
Was the external battery pack replaced? Yes or No
6. O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 998 psi Regulator: 1000 psi
Display and Regulator \pm 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP3002
Lot #: 2023D1E
Exp. Date: 1/1/25
 Results \pm 0.005 of known AC

- 2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.400 AC
 Sim. Ser #: MP6036
 Lot #: 202212C
 Exp. Date: 08Dec24
 Results \pm 5% AC of known AC
- 3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 14323080A4
 Cylinder #: 13
 Exp. Date: 6/5/25
- 4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR7344
 Lot #: ICS 8
 Exp. Date: N/A
 Display reads "Interferent Detect"
- 5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
- 6. Dry Calibration Check (Level 1,C)

Test 1 <u>0.082</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.083</u>
Test 3 <u>0.082</u>	Test 6 <u>0.083</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No
 If No, state reason(s) why: _____

If Yes, change location code back to A.4.

[Signature]
 Inspector Signature

15 Apr 2024
 Date

[Signature]
 Reviewer

16 Apr 2024
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/15/2024 13:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:28
02 Std. Sol.	0.021	13:32
03 Room Air	0.000	13:32
04 Std. Sol.	0.020	13:33
05 Room Air	0.000	13:33
06 Std. Sol.	0.021	13:34
07 Room Air	0.000	13:35


08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 20201E

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC CHECK - 0.020AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/15/2024 13:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:38
02 Std. Sol.	0.402	13:39
03 Room Air	0.000	13:39
04 Std. Sol.	0.404	13:40
05 Room Air	0.000	13:41
06 Std. Sol.	0.403	13:41
07 Room Air	0.000	13:42


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC check - 0.400AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/15/2024 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:54
02 Std. Gas	0.082	13:55
03 Room Air	0.000	13:55
04 Std. Gas	0.082	13:56
05 Room Air	0.000	13:56
06 Std. Gas	0.082	13:57
07 Room Air	0.000	13:57

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/15/2024 13:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:58
02 Std. Gas	0.082	13:58
03 Room Air	0.000	13:59
04 Std. Gas	0.082	13:59
05 Room Air	0.000	14:00
06 Std. Gas	0.083	14:00
07 Room Air	0.000	14:01

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check - #44

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.14.00 09/16
04/15/2024 14:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:06
02 Std. Gas	0.082	14:06
03 Room Air	0.000	14:07
04 Std. Gas	0.083	14:07
05 Room Air	0.000	14:07
06 Std. Gas	0.082	14:08
07 Room Air	0.000	14:08

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000