



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4938 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 14323080A4 Cyl. # 42)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: KDPD
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
 - Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
 - Display: 597 psi Regulator: 600 psi
 - Display and Regulator \pm 50 psi of each other Yes or No
 - Gas tank tare necessary? Yes or No
 - If Yes, display readings after tare (Level 3,M,C,G):
 - Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
 - Wet Calibration Check - Low AC (Level 1,C)
 - Known Value \leq 0.03 AC: 0.020 AC
 - Sim. Ser #: MP3002
 - Lot #: 202301E
 - Exp. Date: 17Jan25
 - Results \pm 0.005 of known AC

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.400 AC
 Sim. Ser #: MP6036
 Lot #: 202212C
 Exp. Date: 08Dec24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 01923080A3
 Cylinder #: 18
 Exp. Date: 2/5/25
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DRT344
 Lot #: ICS 8
 Exp. Date: N/A
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No
 If No, state reason(s) why: _____

If Yes, change location code back to A.4.

[Signature]
 Inspector Signature

16 Apr 2024
 Date

[Signature]
 Reviewer

18 Apr 2024
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/16/2024 07:08

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 42
Standard Expiration? 06/05/2025
Oper No? 133237

Flow Cal. Date: 06/10/2020
Slope 587
Intercept -338327

IR Calibration Date: 06/10/2020

	3um	9um
0th Coef(*100):	-22434	-23179
1st Coef(*100):	269849	136790
2nd Coef(*100):	963	840
H2O adj(mg/l*10k):	355	370

***** Printer Test End *****

Operator Signature
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/16/2024 07:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:10
02 Std. Sol.	0.020	07:17
03 Room Air	0.000	07:17
04 Std. Sol.	0.022	07:18
05 Room Air	0.000	07:18
06 Std. Sol.	0.021	07:19
07 Room Air	0.000	07:20


08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Low Ac check - 0.020Ac*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/16/2024 07:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:21
02 Std. Sol.	0.394	07:22
03 Room Air	0.000	07:23
04 Std. Sol.	0.394	07:24
05 Room Air	0.000	07:24
06 Std. Sol.	0.394	07:25
07 Room Air	0.000	07:25


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac check - 0.400Ac

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/16/2024 07:27

Test	AC	Time
01 Room Air	0.000	07:27
02 *Subject Test	INT*	07:28
03 Room Air	0.000	07:28

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 01923080A3
Cyl No = 18
Expiration Date = 02/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/16/2024 07:29

Test	AC	Time
01 Diagnostic	OK	07:29
02 Room Air	0.000	07:30
03 *Subject Test	RFI*	07:30
04 Room Air	0.000	07:31

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 01923080A3
Cyl No = 18
Expiration Date = 02/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI Check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/16/2024 07:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:32
02 Std. Gas	0.080	07:32
03 Room Air	0.000	07:32
04 Std. Gas	0.080	07:33
05 Room Air	0.000	07:33
06 Std. Gas	0.080	07:34
07 Room Air	0.000	07:34

Lot No = 01923080A3
Cyl No = 18
Exp Date = 02/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check #1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/16/2024 07:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:36
02 Std. Gas	0.080	07:36
03 Room Air	0.000	07:36
04 Std. Gas	0.080	07:37
05 Room Air	0.000	07:37
06 Std. Gas	0.080	07:38
07 Room Air	0.000	07:38

Lot No = 01923080A3
Cyl No = 18
Exp Date = 02/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check #4-6*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.14.00 09/16
04/16/2024 07:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:39
02 Std. Gas	0.080	07:39
03 Room Air	0.000	07:40
04 Std. Gas	0.080	07:40
05 Room Air	0.000	07:41
06 Std. Gas	0.080	07:41
07 Room Air	0.000	07:41

Lot No = 01923080A3
Cyl No = 18
Exp Date = 02/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000