



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4189

A. General Setup and Checks:

1. Initial start up
 - Power On
 - 5-minute countdown during initial warm-up
 - 1-minute countdown to begin test after standby mode
2. Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:
 - Level 1 (Operator)
 - Level 2 (Field Inspector)
 - Level 3 (Lab Personnel)
4. Printer Setup (Level 3,S,P)
5. Forms Load Completed (CMI, Inc., CoBrA Software)

B. Level 1 Setup and Checks:

1. Internal Printer Test (Level 1,P). Sign and attach record.
 - Paper Feed
2. External Printer Test (Level 1,P). Sign and attach record.
3. Barcode Reader (Level 1,B) (Tone sounds after each)
 - Operator ID
 - Invalid ID
 - ND License
 - R key: Refusal. Sign and attach record.

C. Level 2 Setup and Checks:

1. Continuous Air Blank (Level 2,A)
2. Diagnostic (Level 2,D). Sign and attach test record.
3. Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):
 - Inst Ser#: 80-004189 DSP Ver: 1362.14
 - Core Ver: 8164.14 CPLD Ver: 7

D. Level 3 Setup and Checks:

1. DVM Monitor (Level 3,D,D):

3: 12608 s: 0 C: 47.00

9: 13159 s: 1 B: 45

DVM (3 & 9) between 9,000-16,000 counts

RFI indicator with keyed radio: 071

2. Function (F) Keys:

F5 – Tone

F9 - Fan

F6 – Green LED

F10 – Pump

F7 – Red LED

F11 – Solenoid

3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 946 CF: 1.07

4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 210 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T):

C: 47.0 A: 48.5 I: 0.0

B: 45.1 V: 0.0

Cell Temperature (C) between 47 ± 0.02 °C

Breath Hose Temperature (B) between 35.5 – 47 °C

Internal Ambient Temperature (A) above 41 °C

6. Tank Monitor (Level 3,D,G):

Display: 671 psi Regulator: 675 psi

Display and Regulator ± 50 psi of each other Yes or No

Gas tank tare necessary? Yes or No

If Yes, display readings after tare (Level 3,M,C,G):

Display: _____ psi Regulator: _____ psi

7. General Setup (Level 3,S,G)

8. Configure Start Test (Level 3,S,T)

9. Network (Level 3,S,N)

E. Tests (Sign and attach all test records):

1. Configure wet bath simulator for the following test (Level 1,S)

Zero AC-Wet Calibration Check (Level 1,C):

Sim ser #: MP5321 Lot #: NA Exp date: N/A

Known value: 0.000 AC

Results are Zero AC

2. Configure wet bath simulator for the following test (Level 1,S)

Low AC-Wet Calibration Check (Level 1,C):

Known Value ≤ 0.03 AC: 0.020 AC

Sim ser #: MP3002 Lot #: 202301E Exp date: 17Jan25


Results ± 0.005 of known AC

3. Configure wet bath simulator for the following test (Level 1,S)
 High AC-Wet Calibration Check (Level 1,C):
 Known Value \geq 0.25 AC: 0.040 AC
 Sim ser #: MP0036 Lot #: ICS8 Exp date: NIA
 Results \pm 5% of known AC
4. Configure dry gas standard for the following tests (Level 1,S)
 Known Value: 0.080 AC
 Gas Cylinder Lot #: 28423080A3 Cylinder #: 11 Exp. Date: 11/5/25
5. 0.10% Ethanol + 0.05% Acetone (Level 1,B):
 Sim. Ser #: DR7344 Lot #: ICS8 Exp. Date: NIA
 Display reads "Interferent Detect"
6. 0.05% Acetone (Level 1,B):
 Sim. Ser #: DR7344 Lot #: ACE3 Exp. Date: NIA
 Display reads "Interferent Detect"
7. 0.10% Methanol (Level 1,B):
 Sim. Ser #: DR7345 Lot #: MET3 Exp. Date: NIA
 Display Reads "Interferent Detect"
8. 0.10% Isopropanol (Level 1,B):
 Sim. Ser #: DR7350 Lot #: ISO3 Exp. Date: NIA
 Display Reads "Interferent Detect"
9. Ambient Fail (Level 1,B): Display read and test record correct.
10. Invalid Sample (Level 1,B): Display read and test record correct.
11. Sequence Aborted (Level 1,B): Display read test record correct.
12. ABA (Level 1,B): Test record correct.
13. Calibration check out of tolerance (CMS mode): Display read and test record correct.
14. RFI Detected (B,C, or CMS mode): Display read and test record correct.
15. Tank Pressure Below Minimum (CMS mode): Display reads correct.
16. Deficient Samples (CMS mode): Display read and test record correct.
17. Improper Sample (CMS mode): Display read and test record correct.
18. Difference Too Great (CMS mode): Display read and test record correct.
19. CMS (CMS mode): Test record correct.
20. Stability Test (Level 3,T,S):
 Ave: 0.0814 Std Dev: 0.0006 Rel Std Dev: 0.73%

Maintenance/Repairs/Notes: NIA

Instrument is acceptable to be used in the field.
If No, state reason(s) why:

Yes or No



Inspector Signature

17Apr2024

Date



Reviewer

18Apr2024

Date

Intoxilyzer Test Record and Checklist
NDORG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:01

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MemFull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 28423080A3
Standard Cyl #? 37
Standard Expiration? 11/05/2025
Oper No? 133237

Flow Cal. Date: 02/02/2011
Slope 628
Intercept -452608

IR Calibration Date: 05/19/2016
3um 9um

0th Coef(*100):	-17767	-18761
1st Coef(*100):	263439	133606
2nd Coef(*100):	1960	1140
H2O adj(mg/l*10K):	341	362

***** Test End *****



Operator Signature
ANNA NAREHOOD

Remarks:

Internal Print
Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:05

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%\$^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%\$^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 28423080A3
Standard Cyl #? 37
Standard Expiration? 11/05/2025
Oper No? 133237

Flow Cal. Date: 02/02/2011
Slope 628
Intercept -452608

IR Calibration Date: 05/19/2016
 3um 9um

	3um	9um
0th Coef(*100):	-17767	-18761
1st Coef(*100):	263439	133606
2nd Coef(*100):	1960	1140
H2O adj(mg/l*10k):	341	362

***** Printer Test End *****


Operator Signature
ANNA NAREHOOD

Remarks: External Print Test


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:05

Test	AC	Time
01 Room Air	0.000	08:06
02 *Subject Test	REF*	08:06
03 Room Air	0.000	08:07

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = REFUSAL
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 37
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Refusal*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:10

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Diagnostic*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:19
02 Std. Sol.	0.000	08:20
03 Room Air	0.000	08:21
04 Std. Sol.	0.000	08:21
05 Room Air	0.000	08:22
06 Std. Sol.	0.000	08:23
07 Room Air	0.000	08:23


08 Sim Temp = 34.0°C

Simul Ser No = MP5321

Std Sol No = NA

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *zero ac check*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:25
02 Std. Sol.	0.022	08:26
03 Room Air	0.000	08:27
04 Std. Sol.	0.022	08:27
05 Room Air	0.000	08:28
06 Std. Sol.	0.021	08:28
07 Room Air	0.000	08:29

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = 202301E
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Low Ac check - 0.020 Ac*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:31
02 Std. Sol.	0.398	08:32
03 Room Air	0.000	08:32
04 Std. Sol.	0.399	08:33
05 Room Air	0.000	08:34
06 Std. Sol.	0.398	08:34
07 Room Air	0.000	08:35

08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High Ac Check - 0.400AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:44

Test	AC	Time
01 Room Air	0.000	08:44
02 *Subject Test	INT*	08:45
03 Room Air	0.000	08:46

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = INTERFERENT ICS
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Interferent check*
0.10% Ethanol + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:46

Test	AC	Time
01 Room Air	0.000	08:47
02 *Subject Test	INT*	08:47
03 Room Air	0.000	08:48

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = INTERFERENT ACE
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:48

Test	AC	Time
01 Room Air	0.000	08:49
02 *Subject Test	INT*	08:49
03 Room Air	0.000	08:50

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = INTERFERENT MET
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.10% Methanol

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:50

Test	AC	Time
01 Room Air	0.000	08:51
02 *Subject Test	INT*	08:52
03 Room Air	0.000	08:52

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = INTERFERENT ISO
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent check
0.10% Isopropanol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 08:53

Test	AC	Time
01 Room Air	AMB*	08:53
02 Room Air	0.000	08:54

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = AMBIENT
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *ambient fail*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:05

Test	AC	Time
01 Room Air	0.000	09:05
02 Invalid Sample	X.XXX	09:06
03 Room Air	0.000	09:06

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = INVALID
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *invalid sample*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:07

Test	AC	Time
01 Room Air	0.000	09:08
02 *Subject Test	ABT*	09:08
03 Room Air	0.000	09:09

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = ABORTED SEQ
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Sequence aborted*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:10

Test	AC	Time
01 Room Air	0.000	09:11
02 Subject Test 1	0.000	09:11
03 Room Air	0.000	09:11
04 Reported AC	0.000	09:11

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = ABA
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *aba test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:12

Test	AC	Time
01 Diagnostic	OK	09:13
02 Room Air	0.000	09:13
03 Subject Test 1	0.000	09:13
04 Room Air	0.000	09:16
05 Std. Gas	0.017*	09:17
06 Room Air	0.000	09:17

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = CAL CHECK OOT
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: Calibration check
out of tolerance

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	09:18
02 Room Air	0.000	09:18

*Invalid Test
Inhibited - RFI

Lot No = 28423080A3
Cyl No = 11
Exp Date = 11/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:19

Test	AC	Time
01 Diagnostic	OK	09:19
02 Room Air	0.000	09:20
03 *Subject Test	0.000*	09:23
04 Room Air	0.000	09:26
05 Std. Gas	0.080	09:27
06 Room Air	0.000	09:28
07 *Subject Test	0.000*	09:31
08 Room Air	0.000	09:32


No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = DEFICIENT
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: Deficient sample

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:33

Test	AC	Time
01 Diagnostic	OK	09:34
02 Room Air	IPS*	09:34
03 Room Air	0.000	09:35

*Improper Sample

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = IMPROPER
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: *Improper Sample*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:38

Test	AC	Time
01 Diagnostic	OK	09:38
02 Room Air	0.000	09:39
03 Subject Test 1	0.103*	09:39
04 Room Air	0.000	09:42
05 Std. Gas	0.080	09:43
06 Room Air	0.000	09:44
07 Subject Test 2	0.000*	09:44
08 Room Air	0.000	09:45

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: *Difference Too great*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:46


Test	AC	Time
01 Diagnostic	OK	09:46
02 Room Air	0.000	09:47
03 Subject Test 1	0.000	09:47
04 Room Air	0.000	09:49
05 Std. Gas	0.077	09:51
06 Room Air	0.000	09:52
07 Subject Test 2	0.000	09:53
08 Room Air	0.000	09:54
09 Reported AC	0.000	09:47

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = OTH Cit = CMS
Dr. Lic. = ND/DIS821456
Lot No = 28423080A3
Cyl No = 11
Expiration Date = 11/05/2025
County = 08 Oper No. = 133237

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: CMS test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/17/2024 09:55

Test	AC	Time
01 Room Air	0.000	09:56
02 Std. Gas	0.080	09:56
03 Room Air	0.000	09:57
04 Std. Gas	0.081	09:57
05 Room Air	0.000	09:57
06 Std. Gas	0.081	09:58
07 Room Air	0.000	09:58
08 Std. Gas	0.081	09:59
09 Room Air	0.000	09:59
10 Std. Gas	0.081	10:00
11 Room Air	0.000	10:00
12 Std. Gas	0.081	10:00
13 Room Air	0.000	10:01
14 Std. Gas	0.082	10:01
15 Room Air	0.000	10:02
16 Std. Gas	0.082	10:02
17 Room Air	0.000	10:03
18 Std. Gas	0.082	10:03
19 Room Air	0.000	10:04
20 Std. Gas	0.082	10:04
21 Room Air	0.000	10:04
22 Std. Gas	0.081	10:05
23 Room Air	0.000	10:05
24 Std. Gas	0.081	10:06
25 Room Air	0.000	10:06
26 Std. Gas	0.081	10:06
27 Room Air	0.000	10:07
28 Std. Gas	0.082	10:07
29 Room Air	0.000	10:08
30 Std. Gas	0.081	10:08
31 Room Air	0.000	10:09
32 Std. Gas	0.081	10:09
33 Room Air	0.000	10:10
34 Std. Gas	0.081	10:10
35 Room Air	0.000	10:10
36 Std. Gas	0.082	10:11
37 Room Air	0.000	10:11
38 Std. Gas	0.082	10:12
39 Room Air	0.000	10:12
40 Std. Gas	0.082	10:13
41 Room Air	0.000	10:13
42 Cal Check Stats:		
43 Average	0.0814	
44 Std Dev	0.0006	
45 Rel Std Dev	0.73%	



Operator Signature
ANNA NAREHOOD

Remarks:

Stability Test

Form 106-I8000