



**NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION**

INTOXILYZER® 8000 CALIBRATION ADJUSTMENT

Intoxilyzer® 8000 Serial Number: 80-00 4199 Calibration Adjustment Location: TOXL

A. Pre-Adjustment

Replaced Simulator Return O-Ring Yes or No

B. Calibration Adjustment (Level 3,M,C,O)

1. Autocalibration Printout Attached
 - Max Power Res Value ≥ 10
 - Auto Range Res Value ≥ 4
2. Simulator Solutions for Calibration Adjustment

Soln.	g/210 L	Lot No.	Exp. Date	Simulator SN
1	0.000	NA-Milli-Q H ₂ O	NA-Milli-Q H ₂ O	MP5321
2	0.040	<u>202303H</u>	<u>28 Mar 25</u>	<u>MP5289</u>
3	0.080	<u>202302B</u>	<u>14 Feb 25</u>	<u>MP3067</u>
4	0.100	<u>202304A</u>	<u>04 Apr 25</u>	<u>MP6038</u>
5	0.300	<u>202402C</u>	<u>14 Feb 26</u>	<u>MP3062</u>

3. 0.080 AC Calibration Gas for H₂O Adjustment

Lot No. 14323080A4 Cyl No. 43 Exp. Date: 6/5/25

4. Atmospheric Pressure

Displayed by Intoxilyzer® 8000 932 mbar
 Adjusted to using barometer 959 mbar
 Auto Calibration Report printout 959 mbar
 Barometer Model 10510-922
 Barometer Serial Number 230307250
 Barometer Calibration Expiration Date 02 May 25

5. Screen displayed "Calibration Success"

6. Calibration Adjustment Printout Attached

- Solution 1 Avg % Abs ≤ 0.2500
- Solution 2-5 REL STD DEV ≤ 3.000
- Residual (g/210 L) values for solutions 1 - 5 ≤ 0.0020 for 3 μ m and 9 μ m channels

Dry Gas H₂O adjustment sum for 3 μm and 9 μm channels within ± 10
3 μm 3452 (Ave.) + 357 (H₂O Adj.) = 3809
9 μm 3246 (Ave.) + 563 (H₂O Adj.) = 3809

C. Is an Annual Inspection due for this instrument? Yes or No
If Yes, complete Intoxilyzer 8000 Annual Inspection (Document ID: 11698)
If No, complete Intoxilyzer 8000 Calibration (Document ID: 11871).

Remarks/Notes: NIA


Analyst Signature

18 Apr 2024
Date


Reviewer Signature

18 Apr 2024
Date

TOXL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004199
04/18/2024 14:42:11

Auto Calibration
Max Power Res Value = 41
Auto Range Res Value = 18

TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004199
 04/18/2024 14:42:11

Auto Calibration

pg 1 of 2

```

  <<<<<      3um      >>>>>          <<<<<      9um      >>>>>
  -----
  Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1
  Sample      % Abs      (% Abs Ref)      % Abs      (% Abs Ref)
  Sample #1   0.1220      (-0.0120)        0.2270      (-0.0310)
  Sample #2   0.0840      (0.0710)         0.1850      (0.0270)
  Sample #3   0.0810      (0.1290)         0.1980      (0.0470)
  Sample #4   0.0940      (0.1510)         0.2210      (0.0480)
  Avg % Abs   0.0863      (0.1170)         0.2013      (0.0407)
  STD DEV     0.0068      (0.0413)         0.0182      (0.0118)
  REL STD DEV 7.884      (35.323)         9.055       (29.130)
  -----
  
```

```

  Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1
  Sample      % Abs      (% Abs Ref)      % Abs      (% Abs Ref)
  Sample #1   0.8280      (-0.0030)        1.6920      (0.0080)
  Sample #2   0.8190      (0.0190)         1.6780      (0.0160)
  Sample #3   0.7870      (0.0460)         1.6590      (0.0350)
  Sample #4   0.8130      (0.0440)         1.6590      (0.0330)
  Avg % Abs   0.8063      (0.0363)         1.6653      (0.0280)
  STD DEV     0.0170      (0.0150)         0.0110      (0.0104)
  REL STD DEV 2.110      (41.407)         0.659       (37.287)
  -----
  
```

```

  Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
  Sample      % Abs      (% Abs Ref)      % Abs      (% Abs Ref)
  Sample #1   1.4930      (0.0000)         3.0370      (-0.0030)
  Sample #2   1.4880      (0.0290)         3.0420      (0.0050)
  Sample #3   1.4960      (0.0290)         3.0230      (0.0240)
  Sample #4   1.4910      (0.0420)         3.0230      (0.0210)
  Avg % Abs   1.4917      (0.0333)         3.0293      (0.0167)
  STD DEV     0.0040      (0.0075)         0.0110      (0.0102)
  REL STD DEV 0.271      (22.517)         0.362       (61.286)
  -----
  
```

```

  Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1
  Sample      % Abs      (% Abs Ref)      % Abs      (% Abs Ref)
  Sample #1   1.7930      (0.0040)         3.6770      (-0.0130)
  Sample #2   1.8040      (0.0360)         3.6630      (0.0170)
  Sample #3   1.8100      (0.0350)         3.6930      (0.0200)
  Sample #4   1.8490      (0.0260)         3.6950      (0.0180)
  Avg % Abs   1.8210      (0.0323)         3.6837      (0.0183)
  STD DEV     0.0244      (0.0055)         0.0179      (0.0015)
  REL STD DEV 1.342      (17.034)         0.487       (8.332)
  -----
  
```

```

  Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1
  Sample      % Abs      (% Abs Ref)      % Abs      (% Abs Ref)
  Sample #1   5.1670      (-0.0240)        10.0470     (-0.0140)
  Sample #2   5.1630      (0.0140)         10.0660     (0.0190)
  Sample #3   5.1670      (0.0080)         10.0610     (0.0240)
  Sample #4   5.1660      (0.0090)         10.0220     (0.0220)
  Avg % Abs   5.1653      (0.0103)         10.0497     (0.0217)
  STD DEV     0.0021      (0.0032)         0.0241      (0.0025)
  REL STD DEV 0.040      (31.109)         0.240       (11.615)
  -----
  
```

TOXL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-004199
 04/18/2024 14:42:11

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		
-----			-----		
Zero Order Coef	-250.88			-282.96	
First Order Coef	2695.37			1311.83	
Second Order Coef	23.08			13.73	
-----			-----		
Act	Fit	Residual	Act	Fit	Residual
(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)
0.000	-0.000	0.0004	0.000	-0.000	0.0004
0.040	0.041	-0.0007	0.040	0.041	-0.0007
0.080	0.080	-0.0002	0.080	0.080	-0.0002
0.100	0.099	0.0006	0.100	0.099	0.0006
0.300	0.300	-0.0000	0.300	0.300	-0.0000
-----			-----		

<<<<< 3um >>>>>		<<<<< 9um >>>>>	
-----		-----	
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1			
Sample			
Sample #1	3467.00		3286.00
Sample #2	3460.00		3238.00
Sample #3	3430.00		3249.00
Sample #4	3468.00		3253.00
Avg	3452.6667		3246.6667
STD DEV	20.0333		7.7675
REL STD DEV	0.580		0.239
H2O adjust (mg/l*10k)	357		563

Atmospheric Pressure = 959

*****CALIBRATION SUCCESSFUL*****