







Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = STEE      8164.14.00 09/16  
04/18/2024      11:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:13
02 Std. Gas	0.082	11:13
03 Room Air	RFI*	11:14
04 Room Air	0.000	11:14

\*Invalid Test  
Inhibited - RFI

Lot No = 01923080A3  
Cyl No = 60  
Exp Date = 02/05/2025  
County = 22      Oper No. = 133202

  
\_\_\_\_\_  
Operator Signature  
PAIGE SWANSON

Remarks:

Form 106-I8000