



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 5952 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or No (If Yes, Lot # 28423080A3 Cyl. # 45)
  - Keys  Yes or No
  - Power Cord  Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: WISH
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone:  CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 707 psi Regulator: 740 psi  
Display and Regulator ± 50 psi of each other  Yes or No  
Gas tank tare necessary?  Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP3002  
Lot #: 202301E  
Exp. Date: 1/21/25  
 Results ± 0.005 of known AC

2.  Configure simulator for the following test (Level 1,S).  
 Wet Calibration Check - High AC (Level 1,C)  
 Known Value  $\geq$  0.25 AC: 0.400 AC  
 Sim. Ser #: MP6036  
 Lot #: 202212C  
 Exp. Date: 08 Dec 24  
 Results  $\pm$  5% AC of known AC
3.  Configure dry gas standard for the remaining tests (Level 1,S).  
 Known Value : 0.080 AC  
 Gas Cylinder Lot #: 01923080A3  
 Cylinder #: 55  
 Exp. Date: 2/5/25
4.  Interferent Check (Level 1,B)  
 Known Value: 0.10 AC + 0.05% Acetone  
 Sim. Ser #: DR7344  
 Lot #: ICS 8  
 Exp. Date: N/A  
 Display reads "Interferent Detect"
5.  RFI Check (CMS Mode)  
 Display reads "RFI Detect"
6.  Dry Calibration Check (Level 1,C)  

Test 1 <u>0.081</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.082</u>	Test 6 <u>0.081</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

 Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

---



---

Instrument is acceptable to be used in the field.  Yes or No  
 If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4.

[Signature]  
 Inspector Signature

19 Apr 2024  
 Date

[Signature]  
 Reviewer

25 Apr 2024  
 Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                8164.14.00 09/16  
04/19/2024                        08:48

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            28423080A3  
Standard Cyl #?            45  
Standard Expiration?       11/05/2025  
Oper No?                    133237

Flow Cal. Date:            04/16/2021  
Slope                        657  
Intercept                  -744615

IR Calibration Date:        04/16/2021  
                                  3um                        9um

-----  
0th Coef(\*100):            -24450                    -14639  
1st Coef(\*100):            266154                    132500  
2nd Coef(\*100):            2284                      1091  
H2O adj(mg/l\*10k):        516                        323

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA NAREHOOD

Remarks: *Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2024      09:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:24
02 Std. Sol.	0.019	09:25
03 Room Air	0.000	09:25
04 Std. Sol.	0.019	09:26
05 Room Air	0.000	09:26
06 Std. Sol.	0.020	09:27
07 Room Air	0.000	09:28

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *LOW AC check - 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2024      09:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:29
02 Std. Sol.	0.386	09:30
03 Room Air	0.000	09:30
04 Std. Sol.	0.389	09:31
05 Room Air	0.000	09:32
06 Std. Sol.	0.389	09:32
07 Room Air	0.000	09:33


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.400AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                 8164.14.00 09/16  
04/19/2024                        09:34

Test	AC	Time
01 Room Air	0.000	09:34
02 *Subject Test	INT*	09:35
03 Room Air	0.000	09:36

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                     Weight = NA  
Test = OTH                         Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 01923080A3  
Cyl No = 55  
Expiration Date = 02/05/2025  
County = 08                         Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005952  
Location = TOXL                    8164.14.00 09/16  
04/19/2024                            09:36

Test	AC	Time
01 Diagnostic	OK	09:37
02 Room Air	0.000	09:37
03 *Subject Test	RFI*	09:38
04 Room Air	0.000	09:38

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                            Weight = NA  
Test = OTH                                Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 01923080A3  
Cyl No = 55  
Expiration Date = 02/05/2025  
County = 08                                Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2024      09:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:39
02 Std. Gas	0.081	09:40
03 Room Air	0.000	09:40
04 Std. Gas	0.082	09:41
05 Room Air	0.000	09:41
06 Std. Gas	0.082	09:41
07 Room Air	0.000	09:42

Lot No = 01923080A3  
Cyl No = 55  
Exp Date = 02/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Dry cal check #1-3*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2024      09:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:43
02 Std. Gas	0.082	09:43
03 Room Air	0.000	09:44
04 Std. Gas	0.081	09:44
05 Room Air	0.000	09:45
06 Std. Gas	0.081	09:45
07 Room Air	0.000	09:46

Lot No = 01923080A3  
Cyl No = 55  
Exp Date = 02/05/2025  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005952  
Location = TOXL      8164.14.00 09/16  
04/19/2024      09:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:48
02 Std. Gas	0.082	09:48
03 Room Air	0.000	09:49
04 Std. Gas	0.082	09:49
05 Room Air	0.000	09:50
06 Std. Gas	0.082	09:50
07 Room Air	0.000	09:51

Lot No = 01923080A3  
Cyl No = 55  
Exp Date = 02/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000