



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Allen, Kathryn Marie</i>		
Location <i>Cass County Jail - Machine #2</i>	Intoxilyzer® Serial Number <i>80-004942</i>	
Gas Lot Number <i>61923080A3</i>	Gas Cylinder Number <i>23</i>	Gas Expiration Date <i>2/5/25</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.080 AC
 - B. 0.080 AC
 - C. 0.080 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Kathryn M. Allen #3671</i>	Date <i>4-22-24</i>
Reviewed By (Crime Laboratory Use Only) <i>Aime Nuhel</i>	Date <i>29 April 2024</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = CASS 8164.14.00 09/16
04/22/2024 18:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:07
02 Std. Gas	0.080	18:07
03 Room Air	0.000	18:08
04 Std. Gas	0.080	18:08
05 Room Air	0.000	18:09
06 Std. Gas	0.080	18:09
07 Room Air	0.000	18:10

Lot No = 01923080A3
Cyl No = 23
Exp Date = 02/05/2025
County = 09 Oper No. = 090801


Operator Signature
KATHRYN ALLEN

Remarks:

Good test All in Normal
Ranges!
Form 106-I8000