



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-003059</b>	Instrument Location <b>NEW SALEM, ND</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **UPDATE**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <b>Dora B. Miten</b>	Date <b>22 APR 24</b>
---	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <b>Aime Nahl</b>	Certified Date <b>29 April 2024</b>
---	--

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003059  
Location = NS                        8164.14.00 09/16  
04/22/2024                            01:23

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer (Y/N)?                No  
Display Volume?                      No  
Disable On Memfull?                Yes  
# of Print Copies?                    1  
Select Std (D/W/I)?                Dry  
Standard Value?                      0.080  
Standard Lot #?                      01923080A3  
Standard Cyl #?                      24  
Standard Expiration?                02/05/2025  
Oper No?                              079706

Flow Cal. Date:                      11/28/2018  
Slope                                 589  
Intercept                            -276335

IR Calibration Date:                12/04/2018  
    3um                                9um

-----  
0th Coef(\*100):                    -25330                            -15825  
1st Coef(\*100):                    278286                            137135  
2nd Coef(\*100):                    1992                                1257  
H2O adj(mg/l\*10k):                525                                 321

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DANA MIDDLETON

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = NS      8164.14.00 09/16  
04/22/2024      01:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:24
02 Std. Gas	0.081	01:25
03 Room Air	0.000	01:25
04 Std. Gas	0.081	01:25
05 Room Air	0.000	01:26
06 Std. Gas	0.081	01:26
07 Room Air	0.000	01:27

Lot No = 01923080A3  
Cyl No = 24  
Exp Date = 02/05/2025  
County = 30      Oper No. = 079706



Operator Signature  
DANA MIDDLETON

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003059  
Location = NS                    8164.14.00 09/16  
04/22/2024                        01:28

Test	AC	Time
01 Room Air	RFI*	01:31
02 Room Air	RFI*	01:31

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, RFI RFI  
Sub DOB = 10/10/1990  
Sub Sex = Male                    Weight = 150  
Test = OTH                        Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 01923080A3  
Cyl No = 24  
Expiration Date = 02/05/2025  
County = 30                        Oper No. = 079706



Operator Signature  
DANA MIDDLETON

Remarks:

Form 106-I8000