



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 10682 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: CRDS
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 496 psi Regulator: 500 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002
Lot #: 202301E
Exp. Date: 17Jan25
 Results ± 0.005 of known AC

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.400 AC
 Sim. Ser #: MP6036
 Lot #: 202212C
 Exp. Date: 08Dec24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 14323080A4
 Cylinder #: 13
 Exp. Date: 6/5/25
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR7344
 Lot #: ICS 8
 Exp. Date: N/A
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)

Test 1	<u>0.082</u>	Test 4	<u>0.082</u>	Test 7	<u>0.083</u>
Test 2	<u>0.082</u>	Test 5	<u>0.082</u>	Test 8	<u>0.082</u>
Test 3	<u>0.082</u>	Test 6	<u>0.082</u>	Test 9	<u>0.082</u>
		Average	<u>0.082</u>		

 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4.

[Signature]
 Inspector Signature

05Apr2024
 Date

[Signature]
 Reviewer

08Apr2024
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = TOXL 8164.14.00 09/16
04/05/2024 11:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:25
02 Std. Sol.	0.019	11:26
03 Room Air	0.000	11:26
04 Std. Sol.	0.020	11:27
05 Room Air	0.000	11:27
06 Std. Sol.	0.021	11:28
07 Room Air	0.000	11:29

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = 202301E
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: LOW AC check - 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = TOXL 8164.14.00 09/16
04/05/2024 11:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:31
02 Std. Sol.	0.391	11:32
03 Room Air	0.000	11:32
04 Std. Sol.	0.393	11:33
05 Room Air	0.000	11:33
06 Std. Sol.	0.394	11:34
07 Room Air	0.000	11:35


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: High AC Check - 0.400AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = TOXL 8164.14.00 09/16
04/05/2024 11:36

Test	AC	Time
01 Room Air	0.000	11:37
02 *Subject Test	INT*	11:38
03 Room Air	0.000	11:38

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 13
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Interferent Check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = TOXL 8164.14.00 09/16
04/05/2024 11:39

Test	AC	Time
01 Diagnostic	OK	11:39
02 Room Air	0.000	11:40
03 *Subject Test	RFI*	11:40
04 Room Air	0.000	11:41

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 13
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = TOXL 8164.14.00 09/16
04/05/2024 11:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:42
02 Std. Gas	0.082	11:42
03 Room Air	0.000	11:42
04 Std. Gas	0.082	11:43
05 Room Air	0.000	11:43
06 Std. Gas	0.082	11:44
07 Room Air	0.000	11:44

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = TOXL 8164.14.00 09/16
04/05/2024 11:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:45
02 Std. Gas	0.082	11:46
03 Room Air	0.000	11:46
04 Std. Gas	0.082	11:46
05 Room Air	0.000	11:47
06 Std. Gas	0.082	11:47
07 Room Air	0.000	11:48

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks:

Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = TOXL 8164.14.00 09/16
04/05/2024 11:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:49
02 Std. Gas	0.083	11:49
03 Room Air	0.000	11:50
04 Std. Gas	0.082	11:50
05 Room Air	0.000	11:50
06 Std. Gas	0.082	11:51
07 Room Air	0.000	11:51

Lot No = 14323080A4
Cyl No = 13
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check #7-9*

Form 106-I8000