



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006494	Instrument Location Pembina County Sheriff's Department, Cavalier
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) Annual Inspection at Crime Laboratory	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>L. C. Oh</i>	Date 04/05/2024
--	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Ame Nahl</i>	Certified Date 10May2024
--	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = CAVA 8164.14.00 09/16
04/05/2024 21:55

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14323080A4
Standard Cyl #? 6
Standard Expiration? 06/05/2025
Oper No? 131669

Flow Cal. Date: 08/18/2015
Slope 655
Intercept -624238

IR Calibration Date: 05/10/2017
3um 9um

0th Coef(*100):	-28969	-21478
1st Coef(*100):	256589	140860
2nd Coef(*100):	2527	1330
H2O adj(mg/l*10k):	655	473

***** Printer Test End *****



Operator Signature
LUKE OLSON

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = CAVA 8164.14.00 09/16
04/05/2024 21:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:57
02 Std. Gas	0.082	21:57
03 Room Air	0.000	21:58
04 Std. Gas	0.082	21:58
05 Room Air	0.000	21:59
06 Std. Gas	0.082	21:59
07 Room Air	0.000	22:00

Lot No = 14323080A4
Cyl No = 6
Exp Date = 06/05/2025
County = 34 Oper No. = 131669


Operator Signature
LUKE OLSON

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = CAVA 8164.14.00 09/16
04/05/2024 22:01

Test	AC	Time
01 Room Air	RFI*	22:03
02 Room Air	RFI*	22:03

*Invalid Test
Inhibited - RFI

Sub Name = TEST, INSTALL INSTALL
Sub DOB = 11/01/2000
Sub Sex = Male Weight = 150
Test = DUI Cit = 1111111111
Dr. Lic. = ND/111111111111
Lot No = 14323080A4
Cyl No = 6
Expiration Date = 06/05/2025
County = 34 Oper No. = 131669



Operator Signature
LUKE OLSON

Remarks:

RFI Check

Form 106-I8000