

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = BOWM 8164.14.00 09/16
04/24/2024 10:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:52
02 Std. Gas	0.079	10:53
03 Room Air	0.000	10:53
04 Std. Gas	0.077	10:54
05 Room Air	RFI*	10:54
06 Room Air	0.000	10:54

*Invalid Test
Inhibited - RFI

Lot No = 28423080A3
Cyl No = 007
Exp Date = 11/05/2025
County = 06 Oper No. = 130015



Operator Signature
FRANK EBERLE

Remarks:

Form 106-I8000