



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Kenton Kossan</i>		
Location <i>BOTT</i>	Intoxilyzer® Serial Number <i>80-006673</i>	
Gas Lot Number <del><i>221503</i></del> <i>0A23Q843</i> <i>KDR</i>	Gas Cylinder Number <i>052</i>	Gas Expiration Date <i>2/5/2025</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. *0.079* AC  
 B. *0.079* AC  
 C. *0.078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Kenton Kossan</i>	Date <i>4/30/2024</i>
Reviewed By (Crime Laboratory Use Only) <i>Aime Nahl</i>	Date <b>06May2024</b>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006673  
Location = BOTT      8164.14.00 09/16  
04/30/2024      15:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.079	15:05
03 Room Air	0.000	15:05
04 Std. Gas	0.079	15:06
05 Room Air	0.000	15:06
06 Std. Gas	0.078	15:07
07 Room Air	0.000	15:07

Lot No = 01923080A3  
Cyl No = 052  
Exp Date = 02/05/2025  
County = 05      Oper No. = 010907

  
\_\_\_\_\_  
Operator Signature  
KENTON KOSSAN

Remarks:

Form 106-I8000