

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = FAPD 8164.14.00 09/16
05/01/2024 09:36

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01923080A3
Standard Cyl #? 23
Standard Expiration? 02/05/2025
Oper No? 130056

Flow Cal. Date: 05/25/2018
Slope 626
Intercept -548004

IR Calibration Date: 04/09/2024
 3um 9um

0th Coef(*100): -22609 -23575
1st Coef(*100): 260445 132718
2nd Coef(*100): 2885 1338
H2O adj(mg/l*10k): 492 488

***** Printer Test End *****

 #499

Operator Signature
PATRICK THOMAS

Remarks: