



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 7098 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:  
 Gas Cylinder  Yes or  No (If Yes, Lot # 28423080A3 Cyl. # 46)  
 Keys Yes or  No  
 Power Cord Yes or  No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: TIOG
5.  Battery Check  
 Was the external battery pack replaced? Yes or  No
6.  O-Rings  
 Replaced Simulator O-Ring  Yes or  No  
 Replaced Breath Tube O-Ring  Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
 Time Zone:  CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
 Display: 749 psi Regulator: 750 psi  
 Display and Regulator ± 50 psi of each other  Yes or  No  
 Gas tank tare necessary?  Yes or  No  
 If Yes, display readings after tare (Level 3,M,C,G):  
 Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
 Wet Calibration Check - Low AC (Level 1,C)  
 Known Value ≤ 0.03 AC: 0.020 AC  
 Sim. Ser #: MP3002  
 Lot #: 202301E  
 Exp. Date: 17Jan25  
 Results ± 0.005 of known AC

2.  Configure simulator for the following test (Level 1,S).  
 Wet Calibration Check - High AC (Level 1,C)  
 Known Value  $\geq$  0.25 AC: 0.400 AC  
 Sim. Ser #: MP6036  
 Lot #: 202212C  
 Exp. Date: 08Dec24  
 Results  $\pm$  5% AC of known AC
3.  Configure dry gas standard for the remaining tests (Level 1,S).  
 Known Value : 0.080 AC  
 Gas Cylinder Lot #: 01923080A3  
 Cylinder #: 55  
 Exp. Date: 215/25
4.  Interferent Check (Level 1,B)  
 Known Value: 0.10 AC + 0.05% Acetone  
 Sim. Ser #: DR7344  
 Lot #: ICS 8  
 Exp. Date: NIA  
 Display reads "Interferent Detect"
5.  RFI Check (CMS Mode)  
 Display reads "RFI Detect"
6.  Dry Calibration Check (Level 1,C)  

Test 1	<u>0.082</u>	Test 4	<u>0.082</u>	Test 7	<u>0.082</u>
Test 2	<u>0.082</u>	Test 5	<u>0.082</u>	Test 8	<u>0.082</u>
Test 3	<u>0.082</u>	Test 6	<u>0.082</u>	Test 9	<u>0.081</u>
Average <u>0.081</u>					

 Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: NIA

---



---

Instrument is acceptable to be used in the field.  Yes or No

If No, state reason(s) why: \_\_\_\_\_

---

If Yes, change location code back to A.4.

*[Signature]*  
 Inspector Signature

08 May 2024  
 Date

*[Signature]*  
 Reviewer

08 May 2024  
 Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007098  
Location = TOXL      8164.16.00 09/18  
05/08/2024      10:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Sol.	0.020	10:24
03 Room Air	0.000	10:24
04 Std. Sol.	0.020	10:25
05 Room Air	0.000	10:26
06 Std. Sol.	0.019	10:26
07 Room Air	0.000	10:27

08 Sim Temp = 34.0°C

Simul Ser No = MP3002  
Std Sol No = 202301E  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: LOWAC check - 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007098  
Location = TOXL      8164.16.00 09/18  
05/08/2024      10:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Sol.	0.393	10:36
03 Room Air	0.000	10:37
04 Std. Sol.	0.394	10:38
05 Room Air	0.000	10:38
06 Std. Sol.	0.396	10:39
07 Room Air	0.000	10:40


08 Sim Temp = 34.0°C

Simul Ser No = MP6036

Std Sol No = 202212C

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High Ac Check - 0.400Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007098  
Location = TOXL      8164.16.00 09/18  
05/08/2024      10:40

Test	AC	Time
01 Room Air	0.000	10:41
02 *Subject Test	INT*	10:41
03 Room Air	0.000	10:42

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 01923080A3  
Cyl No = 55  
Expiration Date = 02/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007098  
Location = TOXL      8164.16.00 09/18  
05/08/2024      10:42

Test	AC	Time
01 Diagnostic	OK	10:43
02 Room Air	0.000	10:44
03 *Subject Test	RFI*	10:44
04 Room Air	0.000	10:44

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 01923080A3  
Cyl No = 55  
Expiration Date = 02/05/2025  
County = 08      Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007098  
Location = TOXL      8164.16.00 09/18  
05/08/2024      10:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:45
02 Std. Gas	0.082	10:46
03 Room Air	0.000	10:46
04 Std. Gas	0.082	10:46
05 Room Air	0.000	10:47
06 Std. Gas	0.082	10:47
07 Room Air	0.000	10:48

Lot No = 01923080A3  
Cyl No = 55  
Exp Date = 02/05/2025  
County = 08      Oper No. = 133237

  
Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007098  
Location = TOXL      8164.16.00 09/18  
05/08/2024      10:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:49
02 Std. Gas	0.082	10:49
03 Room Air	0.000	10:49
04 Std. Gas	0.082	10:50
05 Room Air	0.000	10:50
06 Std. Gas	0.082	10:51
07 Room Air	0.000	10:51

Lot No = 01923080A3  
Cyl No = 55  
Exp Date = 02/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007098  
Location = TOXL      8164.16.00 09/18  
05/08/2024      10:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:52
02 Std. Gas	0.082	10:52
03 Room Air	0.000	10:53
04 Std. Gas	0.082	10:53
05 Room Air	0.000	10:54
06 Std. Gas	0.081	10:54
07 Room Air	0.000	10:55

Lot No = 01923080A3  
Cyl No = 55  
Exp Date = 02/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry Cal Check #7-9

Form 106-I8000