




**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-005946</b>	Instrument Location <b>FINL</b>
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Annual Inspection at Crime Laboratory</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- ~~N/A~~  2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>5/13/2024</b>
--	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By	Certified Date
-----------------------	----------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005946  
Location = FINL                 8164.14.00 09/16  
05/13/2024                        19:10

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 14323080A1  
Standard Cyl #?                 22  
Standard Expiration?            06/05/2025  
Oper No?                        130363

Flow Cal. Date:                 06/02/2021  
Slope                            642  
Intercept                        -600927

IR Calibration Date:            06/02/2021  
                                    3um                            9um

-----  
0th Coef(\*100):                -20165                        -16483  
1st Coef(\*100):                290371                        141330  
2nd Coef(\*100):                953                            910  
H2O adj(mg/l\*10k):            496                            404

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
-----  
Operator Signature  
RICK CZAPIEWSKI

Remarks:                        *Re install after  
Annual inspect*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005946  
Location = FINL      8164.14.00 09/16  
05/13/2024      19:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:11
02 Std. Gas	0.082	19:12
03 Room Air	0.000	19:12
04 Std. Gas	0.081	19:13
05 Room Air	0.000	19:13
06 Std. Gas	0.080	19:14
07 Room Air	0.000	19:14

Lot No = 14323080A1  
Cyl No = 22  
Exp Date = 06/05/2025  
County = 46      Oper No. = 130363



Operator Signature  
RICK CZAPIEWSKI

Remarks:

*Install after  
Annual inspection*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005946  
Location = FINL                    8164.14.00 09/16  
05/13/2024                            19:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	19:16
02 Room Air	0.000	19:16

\*Invalid Test  
Inhibited - RFI

Lot No = 14323080A1  
Cyl No = 22  
Exp Date = 06/05/2025  
County = 46                            Oper No. = 130363



Operator Signature  
RICK CZAPIEWSKI

Remarks:

*RFI detect*

Form 106-I8000