




INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number	80-006504	Instrument Location	CARR
Reason for Install/Repair			
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) Annual Inspection at Crime Laboratory			


Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- AEN 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature		Date	5/30/2024
---------------------------	---	------	-----------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By		Certified Date	31May2024
-----------------------	---	----------------	-----------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = CARR 8164.14.00 09/16
05/30/2024 18:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:12
02 Std. Gas	0.081	18:13
03 Room Air	0.000	18:13
04 Std. Gas	0.081	18:13
05 Room Air	0.000	18:14
06 Std. Gas	0.080	18:14
07 Room Air	0.000	18:15

Lot No = 01923080A3
Cyl No = 25
Exp Date = 02/05/2025
County = 16 Oper No. = 131530


Operator Signature
CHRISTOPHER ORMISTON

Remarks: *Good Test - ACA*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = CARR 8164.14.00 09/16
05/30/2024 18:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	18:20
02 Room Air	RFI*	18:20

*Invalid Test
Inhibited - RFI

Lot No = 01923080A3
Cyl No = 25
Exp Date = 02/05/2025
County = 16 Oper No. = 131530



Operator Signature
CHRISTOPHER ORMISTON

Remarks:

Good Test ~ RFI

Form 106-I8000