



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 3067 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # 01923080A3 Cyl. # 16)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: GFTN
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 791 psi Regulator: 750 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002
Lot #: 202301E
Exp. Date: 1/1/25
 Results ± 0.005 of known AC

2. Configure simulator for the following test (Level 1,S).
 Wet Calibration Check - High AC (Level 1,C)
 Known Value \geq 0.25 AC: 0.400 AC
 Sim. Ser #: MP6036
 Lot #: 202212C
 Exp. Date: 08Dec24
 Results \pm 5% AC of known AC
3. Configure dry gas standard for the remaining tests (Level 1,S).
 Known Value : 0.080 AC
 Gas Cylinder Lot #: 14323080A4
 Cylinder #: 12
 Exp. Date: 6/5/25
4. Interferent Check (Level 1,B)
 Known Value: 0.10 AC + 0.05% Acetone
 Sim. Ser #: DR7344
 Lot #: 1CS8
 Exp. Date: N/A
 Display reads "Interferent Detect"
5. RFI Check (CMS Mode)
 Display reads "RFI Detect"
6. Dry Calibration Check (Level 1,C)

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.080</u>	Test 9	<u>0.081</u>
Average <u>0.080</u>					

 Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No
 If No, state reason(s) why: _____

If Yes, change location code back to A.4.

[Signature]
 Inspector Signature

31 May 2024
 Date

[Signature]
 Reviewer

03 June 2024
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.14.00 09/16
05/31/2024 13:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01923080A3
Standard Cyl #? 16
Standard Expiration? 02/05/2025
Oper No? 133237

Flow Cal. Date: 04/17/2020
Slope 609
Intercept -236750

IR Calibration Date: 04/17/2020
 3um 9um

0th Coef(*100): -19307 -18350
1st Coef(*100): 282277 142540
2nd Coef(*100): 1257 744
H2O adj(mg/l*10k): 302 214

***** Printer Test End *****



Operator Signature
ANNA NAREHOOD

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.14.00 09/16
05/31/2024 13:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:47
02 Std. Sol.	0.020	13:48
03 Room Air	0.000	13:49
04 Std. Sol.	0.020	13:50
05 Room Air	0.000	13:50
06 Std. Sol.	0.019	13:51
07 Room Air	0.000	13:51

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = 202301E
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *LOW AC Check - 0.020AE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.14.00 09/16
05/31/2024 13:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:53
02 Std. Sol.	0.382	13:54
03 Room Air	0.000	13:54
04 Std. Sol.	0.384	13:55
05 Room Air	0.000	13:55
06 Std. Sol.	0.385	13:56
07 Room Air	0.000	13:57

08 Sim Temp = 34.0°C

Simul Ser No = MP6036
Std Sol No = 202212C
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *High AC check - 0.400AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.14.00 09/16
05/31/2024 14:00

Test	AC	Time
01 Diagnostic	OK	14:01
02 Room Air	0.000	14:02
03 *Subject Test	RFI*	14:02
04 Room Air	0.000	14:02

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 14323080A4
Cyl No = 12
Expiration Date = 06/05/2025
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA NAREHOOD

Remarks: *RFI check*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.14.00 09/16
05/31/2024 14:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:03
02 Std. Gas	0.080	14:04
03 Room Air	0.000	14:04
04 Std. Gas	0.081	14:04
05 Room Air	0.000	14:05
06 Std. Gas	0.081	14:05
07 Room Air	0.000	14:06

Lot No = 14323080A4
Cyl No = 12
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: *Dry cal check #1-3*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.14.00 09/16
05/31/2024 14:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:07
02 Std. Gas	0.080	14:07
03 Room Air	0.000	14:08
04 Std. Gas	0.081	14:08
05 Room Air	0.000	14:08
06 Std. Gas	0.080	14:09
07 Room Air	0.000	14:09

Lot No = 14323080A4
Cyl No = 12
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #4-6
4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = TOXL 8164.14.00 09/16
05/31/2024 14:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:10
02 Std. Gas	0.080	14:11
03 Room Air	0.000	14:11
04 Std. Gas	0.081	14:11
05 Room Air	0.000	14:12
06 Std. Gas	0.081	14:12
07 Room Air	0.000	14:13

Lot No = 14323080A4
Cyl No = 12
Exp Date = 06/05/2025
County = 08 Oper No. = 133237



Operator Signature
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000