



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6687 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or No (If Yes, Lot # 14323080A4 Cyl. # 7)
  - Keys  Yes or No
  - Power Cord  Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: PLPD
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring Yes or  No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone:  CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 983 psi Regulator: 975 psi  
Display and Regulator ± 50 psi of each other  Yes or  No  
Gas tank tare necessary? Yes or  No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP3002  
Lot #: 202301E  
Exp. Date: 1/24/25  
 Results ± 0.005 of known AC

2.  Configure simulator for the following test (Level 1,S).  
 Wet Calibration Check - High AC (Level 1,C)  
 Known Value  $\geq$  0.25 AC: 0.400 AC  
 Sim. Ser #: MP6036  
 Lot #: 202212C  
 Exp. Date: 08Dec24  
 Results  $\pm$  5% AC of known AC
3.  Configure dry gas standard for the remaining tests (Level 1,S).  
 Known Value : 0.080 AC  
 Gas Cylinder Lot #: 28423080A3  
 Cylinder #: 11  
 Exp. Date: 11/5/25
4.  Interferent Check (Level 1,B)  
 Known Value: 0.10 AC + 0.05% Acetone  
 Sim. Ser #: DR7344  
 Lot #: ICS 8  
 Exp. Date: NIA  
 Display reads "Interferent Detect"
5.  RFI Check (CMS Mode)  
 Display reads "RFI Detect"
6.  Dry Calibration Check (Level 1,C)  

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
		Average	<u>0.081</u>	<u>0.080</u>	<u>ave</u>

 Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: NIA

---



---

Instrument is acceptable to be used in the field.  Yes or No  
 If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4.

*Ann Miller*  
 Inspector Signature

18 June 2024  
 Date

*Janelle Puschke*  
 Reviewer

19 June 2024  
 Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/18/2024      13:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:02
02 Std. Sol.	0.021	13:03
03 Room Air	0.000	13:04
04 Std. Sol.	0.021	13:04
05 Room Air	0.000	13:05
06 Std. Sol.	0.021	13:06
07 Room Air	0.000	13:06

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 202301E

County = 08

Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *LOW AC check- 0.020Ac*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/18/2024      13:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:14
02 Std. Sol.	0.400	13:15
03 Room Air	0.000	13:15
04 Std. Sol.	0.401	13:16
05 Room Air	0.000	13:17
06 Std. Sol.	0.401	13:17
07 Room Air	0.000	13:18

08 Sim Temp = 34.0°C

Simul Ser No = MP6036  
Std Sol No = 202212C  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High AC Check - 0.400AC

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006687  
Location = TOXL                    8164.14.00 09/16  
06/18/2024                            13:20

Test	AC	Time
01 Room Air	0.000	13:20
02 *Subject Test	INT*	13:21
03 Room Air	0.000	13:22

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                            Weight = NA  
Test = OTH                            Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 28423080A3  
Cyl No = 11  
Expiration Date = 11/05/2025  
County = 08                            Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA NAREHOOD

Remarks: *Interferent check*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/18/2024      13:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:26
02 Std. Gas	0.081	13:26
03 Room Air	0.000	13:27
04 Std. Gas	0.081	13:27
05 Room Air	0.000	13:28
06 Std. Gas	0.081	13:28
07 Room Air	0.000	13:29

Lot No = 28423080A3  
Cyl No = 11  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: *dry cal check #1-3*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/18/2024      13:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.081	13:30
03 Room Air	0.000	13:30
04 Std. Gas	0.080	13:31
05 Room Air	0.000	13:31
06 Std. Gas	0.081	13:31
07 Room Air	0.000	13:32

Lot No = 28423080A3  
Cyl No = 11  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006687  
Location = TOXL      8164.14.00 09/16  
06/18/2024      13:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:33
02 Std. Gas	0.081	13:34
03 Room Air	0.000	13:34
04 Std. Gas	0.081	13:35
05 Room Air	0.000	13:35
06 Std. Gas	0.081	13:35
07 Room Air	0.000	13:36

Lot No = 28423080A3  
Cyl No = 11  
Exp Date = 11/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000