



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-006501</b>	Instrument Location <b>Minot AFB</b>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <b>AEN</b> <input checked="" type="checkbox"/> Other (Specify) <u>Annual Inspection at Crime Laboratory</u>	

Check When Done:

- AEN**  1. Surge Protector Installed/Property Grounded.
- AEN**  2. Telephone Line Connected to Intoxilyzer® 8000.
- AEN**  3. Breath Tube Heated.
- AEN**  4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- AEN**  5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
  - 6. Run Tests:
    - A. Print Test (Level1, Function P).
    - B. ACA Test (Level 1, Function C).
    - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- AEN**  7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- AEN**  8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- AEN**  9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- AEN**  10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>18 June 2024</b>
--	-----------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date <b>18 June 2024</b>
--	---------------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006501  
Location = MNAF                    8164.14.00 09/16  
06/18/2024                            13:34

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            14323080A4  
Standard Cyl #?            5  
Standard Expiration?        06/05/2025  
Oper No?                    133654

Flow Cal. Date:            06/13/2019  
Slope                      651  
Intercept                  -777821

IR Calibration Date:        06/13/2019

	3um	9um
0th Coef(*100):	-20531	-14906
1st Coef(*100):	270252	139749
2nd Coef(*100):	1917	1284
H2O adj(mg/l*10k):	393	321

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
EMILY HENLEY

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006501  
Location = MNAF      8164.14.00 09/16  
06/18/2024      13:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Gas	0.079	13:40
03 Room Air	0.000	13:40
04 Std. Gas	0.079	13:40
05 Room Air	0.000	13:41
06 Std. Gas	0.079	13:41
07 Room Air	0.000	13:42

Lot No = 14323080A4  
Cyl No = 5  
Exp Date = 06/05/2025  
County = 64      Oper No. = 133654



Operator Signature  
EMILY HENLEY

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006501  
Location = MNAF      8164.14.00 09/16  
06/18/2024      13:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:36
02 Std. Gas	0.080	13:36
03 Room Air	0.000	13:37
04 Std. Gas	0.079	13:37
05 Room Air	RFI*	13:38
06 Room Air	RFI*	13:38

\*Invalid Test  
Inhibited - RFI

Lot No = 14323080A4  
Cyl No = 5  
Exp Date = 06/05/2025  
County = 64      Oper No. = 133654



Operator Signature  
EMILY HENLEY

Remarks:

Form 106-I8000