



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6678 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder  Yes or No (If Yes, Lot # 01923080A3 Cyl. # 57)
  - Keys  Yes or No
  - Power Cord  Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: GFPD
5.  Battery Check
  - Was the external battery pack replaced? Yes or  No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or  No
  - Replaced Breath Tube O-Ring  Yes or No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone:  CST or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: 724 psi Regulator: 725 psi  
Display and Regulator ± 50 psi of each other  Yes or No  
Gas tank tare necessary? Yes or  No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP3002  
Lot #: 202301E  
Exp. Date: 1/1/25  
 Results ± 0.005 of known AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006678  
Location = TOXL                8164.14.00 09/16  
06/19/2024                      11:11

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            01923080A3  
Standard Cyl #?            57  
Standard Expiration?       02/05/2025  
Oper No?                    133237

Flow Cal. Date:            08/22/2016  
Slope                        686  
Intercept                   -611596

IR Calibration Date:       08/22/2016  
                                    3um                      9um

-----  
0th Coef(\*100):            -15445                    -21650  
1st Coef(\*100):            251854                   131766  
2nd Coef(\*100):            3094                      1378  
H2O adj(mg/l\*10k):        549                        513

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA NAREHOOD

Remarks:    Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006678  
Location = TOXL      8164.14.00 09/16  
06/19/2024      11:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:26
02 Std. Sol.	0.020	11:27
03 Room Air	0.000	11:27
04 Std. Sol.	0.020	11:28
05 Room Air	0.000	11:28
06 Std. Sol.	0.020	11:29
07 Room Air	0.000	11:30

08 Sim Temp = 34.0°C

Simul Ser No = MP3002  
Std Sol No = 202301E  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Low AC check - 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006678  
Location = TOXL                 8164.14.00 09/16  
06/19/2024                        11:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:37
02 Std. Sol.	0.398	11:37
03 Room Air	0.000	11:38
04 Std. Sol.	0.398	11:39
05 Room Air	0.000	11:39
06 Std. Sol.	0.398	11:40
07 Room Air	0.000	11:41

08 Sim Temp = 34.0°C

Simul Ser No = MP6036  
Std Sol No = 202212C  
County = 08                        Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: High Ae Check - 0.400 Ae

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006678  
Location = TOXL      8164.14.00 09/16  
06/19/2024      11:42

Test	AC	Time
01 Room Air	0.000	11:43
02 *Subject Test	INT*	11:43
03 Room Air	0.000	11:44

\*Invalid Test  
Interferent Detected

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = INTERFERENT CK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A4  
Cyl No = 12  
Expiration Date = 06/05/2025  
County = 08      Oper No. = 133237

  
Operator Signature  
ANNA NAREHOOD

Remarks: Interferent check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006678  
Location = TOXL                              8164.14.00 09/16  
06/19/2024                                              11:46

Test	AC	Time
01 Diagnostic	OK	11:47
02 Room Air	RFI*	11:47
03 Room Air	0.000	11:48

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                                      Weight = NA  
Test = OTH                                              Cit = RFI CHECK  
Dr. Lic. = ND/TES989643  
Lot No = 14323080A4  
Cyl No = 12  
Expiration Date = 06/05/2025  
County = 08                                              Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA NAREHOOD

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006678  
Location = TOXL      8164.14.00 09/16  
06/19/2024      11:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:49
02 Std. Gas	0.080	11:49
03 Room Air	0.000	11:50
04 Std. Gas	0.080	11:50
05 Room Air	0.000	11:51
06 Std. Gas	0.080	11:51
07 Room Air	0.000	11:52

Lot No = 14323080A4  
Cyl No = 12  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #1-3

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006678  
Location = TOXL      8164.14.00 09/16  
06/19/2024      11:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:53
02 Std. Gas	0.079	11:53
03 Room Air	0.000	11:53
04 Std. Gas	0.080	11:54
05 Room Air	0.000	11:54
06 Std. Gas	0.080	11:55
07 Room Air	0.000	11:55

Lot No = 14323080A4  
Cyl No = 12  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check # 4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006678  
Location = TOXL      8164.14.00 09/16  
06/19/2024      11:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Gas	0.081	11:56
03 Room Air	0.000	11:57
04 Std. Gas	0.081	11:57
05 Room Air	0.000	11:58
06 Std. Gas	0.080	11:58
07 Room Air	0.000	11:59

Lot No = 14323080A4  
Cyl No = 12  
Exp Date = 06/05/2025  
County = 08      Oper No. = 133237



Operator Signature  
ANNA NAREHOOD

Remarks: Dry cal check #7-9

Form 106-I8000